Rohera Healthcare & Technology Pvt. Ltd.



CREDENTIALS & PATIENT'S REPORTS

www.emedica.in info@emedica.in Pune | Mumbai | Delhi | Bangalore

1.	About eMedica	Page 03
2.	Ethical Papers by Maharashtra Govt DMER & Arogya Bhawan	Page 05
3.	Clinical Trial Reports by Daffrin Hospital (Maharashtra Govt.)	Page 09
4.	Calibration Report	Page 16
5.	Certificates	Page 27
6.	Recognition	Page 32
7.	Patent Copy	Page 37
8.	Patients Path-Lab Reports	
	I. BSL Trial Reports at Serum Institute	Page 41
	II. Random BSL Reports	Page 42
	III. Random Cholesterol Reports	Page 52
	IV. HbA1C Reports	Page 58
	V. Creatinine Reports	Page 68
	VI. Thyroid Report	Page 76
	VII. Heart Blockage/Condition Reports	Page 78
	VIII.Refractive Error Report	Page 82
	IX Leukaemia Renort	Page 84



The Hypothesis:

The universe is all about energy. No matter whether you call the energy chi, prana, electrons, or some other name. the universe is about interactions of energy.

In general, it is observed that frequencies can repair damage tissues and cells within body. This is a primary basis for the therapeutic use of this technology, stimulating tissues at a cellular level. Electrical pulse and frequencies passing through our whole body will have a positive charge to our trillions of cells.

Basic cell functions and the effect of Electromagnetic Frequencies:

Our skin, bones, and organs are composed of tiny cells. The membrane of a healthy cell has both positive and negative charges that are required for the exchange of potassium, sodium, and calcium ions. When cells become distressed form disease, trauma or toxins, they lose their ability to function efficiently, eMedica restores the positive and negative charge in the cell, enabling it to perform its natural functions while speeding tissues recovery.

Everyone's body respond differently to electron pulse waves and different frequencies target specific tissues types. eMedica emits a unique series of pulsed electro fields at precise frequencies targeted on four key tissues types resulting in increased circulation. Reduced inflammation, improve modality, and relieved pain.

All cells need energy to function through the increased motion of ions and electrolytes magnetic cells help cells increase their energy or charge.

Electro fields also affect the charges of the cell membrane, which allows membrane channels to open up.

Resting cells are negatively charged on the inside, while the outside of the cell is more positively charged. The flow of charge across the cell membrane is what generates electrical current.

When a cell is stimulated, It allows positive charges to enter the cell through open ion channels. The inside of the cell then becomes more positively charged, which trigger further electrical current that can turn into electrical pulses, called action potentials. Our bodies use certain patterns if action potentials to initiate the correct movements, thoughts and behaviour.

3



A disruption in electrical current can lead to illness. eMedica Electro Field therapy can restore the disruption of the electrical current to its normal state, therefore. Helping restore the cells. (Source)

Electric and magnetic fields (EMFs) are invisible areas of energy, often referred to as radiation, that are associated with the use of electrical power and various forms of natural and man- made lighting. EMFs are typically characterized by wavelength or frequency into on of two radioactive categories:

Non-ionizing: Low level radiation which is generally perceived as harmless to humans Ionizing: High – level radiation which was the potential for cellular and DNA damage Along with the increased voltage, the body also need the raw materials (power nutrition) to heal. If your body does not have the energy storehouse necessary to produce healing, the voltage in the cells lessens, and when that happens, the body will become diseased. These channels are like doors and window of a house. By opening cell channels, nutrients are better able to enter the cell and waste is more easily eliminated from the cell. This helps to re balance and restore optimum cell function. If you restore enough cells. They will all work efficiently. Cells of the same types come together to make tissues and those tissues come together to make organs. So, by restoring or maintaining cellular function, allowing the entire body to function better. We all know that body ages over time. Maintaining the function of every individual cell at an optimal level every day is an important part of slowing aging.

Can eMedica therapy protect me from future illnesses?

Magnetic fields protect against cell injury by improving circulation, energy and repair process. Magnetic fields balance cells, tissues and bodily functions at very fundamental levels. A key to understanding Energetic Medicine is to understand that each cell is designed to run at a specific voltage and a specific frequency. Generally speaking, diseases is caused when cells have to little voltage and are running at too low a frequency. To operate correctly, cells must have both the proper voltage and the proper frequency. Chronic illness is almost always characterized by low voltage and a decrease in the frequency of the affected organs.

Restoration of health must involve correcting both the voltage and the frequency of each cell and providing the nutrition necessary to make good new cells.



ETHICAL PAPERS BY GOVERNMENT OF MAHARASHTRA













संचालनालय, वैद्यकीय शिक्षण आणि संशोधन, मुंबई



शासकीय देव प्राविधालय व संग्यालय इमारत चीधा पत्रला, सेंट तांजेयु कामाल्य कावार, पी. तीवेली रोड, पोर्ट, पुंबई - २०० ००६ दुरप्यती - ६९ -२२ - २२६२०३६९ -६७/२०६५२२०१ (५७/५५) स्क्लिस्थळ: Improvement edit in to improvement more

क.संवैशिवसं/संकीर्ण/eMedica/इलेक्टॉनिक मेडिकल गॅंजेट/ १४-३४/२०२१, दिनांक १४/१९/२०२९ कार्यालयीन आदेश

विषय : The innovative Wellness product_eMedica.

संदर्भ : सीईओ, रोहेरा हेल्थकेअर ॲन्ड टेक्नॉलॉजी प्रा.लि, पुणे यांचे पत्र दिनांक १२/१०/२०२१

प्रस्तुत विषयी संदर्भीय पत्रान्वयं सीईओ, रोहेरा हेल्थकेअर ॲन्ड टेक्नॉलॉजी प्रा.लि., पुणे यांनी "eMedica" या उत्पादनाचे शासकीयं रुग्णालयामध्ये ट्रायल करण्यास परवानगी मिळणेबाबत विनंती केली आहे.

उक्त प्रस्तावामध्ये पुढील बाबी नमुद करण्यात आल्या आहेत.

"eMedica" या उपकरणाची उपयुक्तता

Improves Immunity, Support to reduce Heart blockages, Support to control Diabetes, Support to fight different viral infection, Support to control Hypothyroidism, Improves Blood Circulation, Increase Oxygen saturation, Balance the pH level, Control Cholesterol levels, Control Blood Pressure, Support to control Parkinsons & Paralysis, Prevent Cencer & Metastasis, Support to detox liver and kidney, Support to control Uric acid levels (Gout), Support to reduce lung diseases and Fibrosis of lungs. No side effects, Non invasive, Family device.

- उक्त उपकरणास अन्न व औषध प्रशासनाची मान्यता मिळण्याकरीता संबंधित कंपनीमार्फत प्रस्ताव सादर करण्यात आलेला आहे.
- eMedica या उपकरणास NABL Approved, CE, ROSH, ISO Certified असे विविध नामांकीत प्रमाणपत्रे प्राप्त झाली असल्याचे प्रस्तावात नमुद केले आहे.

रोहेरा हेल्थकेअर ॲन्ड टेक्नॉलॉजी प्रा.लि., पुणे यांनी eMedica या इलेक्टॉनिक मेडिकल गॅजेट उपकरणाच्या उपयोगासंदर्भात, अधोरेखित केलेल्या उक्त बाबींच्या पडताळणीसह पुढील नमुद मुद्यांवावत देखिल स्पष्टता होणे आवश्यक आहे.

- सदरचे उपकरण रुग्णोपचाराच्या दृष्टिने उपयुक्त आहे का ? असल्यास प्रचलित उपचार पध्दतीच्या तुलनेत किती प्रमाणात फायदेशिर आहे ?
- शासकीय रुग्णालयात सदर उपकरणाचा वापर करणे शक्य आहे किया कसे ?



यासर्य बाबींचा उहापोह करून सदर उपकरणाचे आयुर्मान, उपयोग सुलभता इत्यादींच्या परिक्षणात्मक अभ्यासांती अहवाल प्राप्त होणे आवश्यक आहे.

यास्तव eMedica या उत्पादनाचे सर्वांगीन चाचणीअंती अहवाल सादर करण्याकरीता पुढील प्रमाणे तज्ञ समिती गठित करण्यात येत आहे.

- १) डॉ. विद्या नगार, प्राध्यापक (तदर्थ), औषधवैद्यकशास्त्र, ग्रॅ.शा.वै.म., मुंबई
- २) डॉ. गिता शेठ, सहयोगी प्राध्यापक, नेफ्रोलॉजी, ग्रॅ.शा.वै.म., मुंबई
- ३) डॉ. तेजल वधान (सुरळकर), सहयोगी प्राध्यापक, कामा व अल्ब्लेस रुग्णालय, मुंबई

उपरोक्त समितीने eMedica या इलेक्टॉनिक मेडिकल गॅजेट उपकरणाचे परिक्षण करुन स्वयंस्पष्ट अहवाल १० दिवसांत संचालनालयास सादर करावा.

संचालक,

वैद्यकीय शिक्षण व संशोधन, मुंबई

प्रति,

बॉ. विद्या नगार,
 प्राध्यापक (तदर्थ), औषधवैद्यकशास्त्र, ग्रॅ.शा.वै.म., मुंबई

२) डॉ. गिता शेठ, सहयोगी प्राध्यापक, नेफ्रोलॉजी, ग्रॅ.शा.वै.म., मुंबई

इ) डॉ. तेजल वधान (सुरळकर),
 सहयोगी प्राध्यापक, कामा व अल्ब्लेस रुग्णालय, मुंबई


्रातः सीईओ, रोहेरा हेल्थकेअर ॲन्ड टेक्नॉलॉजी प्रा.लि, पुणे

यांना कळविण्यात येते की, eMedica या उपकरणाचे उक्त समितीसमोर सादरीकरण करावे. तसेच तपासणी/प्रात्यक्षिकाकरीता उपकरणी उपलब्ध करुन द्यावे.

प्रत: माहितीस्तव.

१) अधिष्ठाता, ग्रॅन्ट शासकीय वैद्यकीय महाविद्यालय, मुंबई

२) अधिक्षक, कामा व अल्ब्लेस रुग्णालय, मुंबई











रुग्णालय (रिक्रियेस्तर), मुंबई

आरोग्य सेंवा आयुक्तालय ,

(महाराष्ट्र राज्य)

आरोग्य भवन, ६वा मजला, सेंट जॉर्जेस रुग्णालय, आवार, पी डिमेलो रोड, मुंबई ४०० ००१

ha-arogya.gov.in nail.com	
969-6	08
	/55
1	/८५/२१/ई-मिडीका/ दि. ५ जानेवारी

प्रति,

मेटस्वतंत्र पॉवर प्रा.लि., २०१, मनिष प्लाझा, भारामा मॅजेस्टीक, एन.आय.बी.एम.रोङ, पुणे ४११ ०४८

विषय - ई-मिडिका बाबत...

संदर्भ - आपले दि.२८.७.२०२१ चे पत्र

उपरोक्त विषयी आपणांस कळविण्यांत येते की, आपल्या संदर्भाधीन पत्रान्वये ई-मिडिका या उत्पादनावायतचे माहितीपत्रक व इतर कागदपत्रे आयुक्तालयास प्राप्त झालेली आहे. सदरहू पत्रान्वये आपल्या कंपनीने ई-मिडिका हे उपकरण मधुमेह, कॉलेस्ट्रोल, हृदयविकार, इ. आजारांवर अत्यंत उपयुक्त असल्यावावत कळिलिले आहे.

तरी आपणांस कळविण्यांत येते की, आपले उपकरण हे विभागीय संदर्भ सेवा रूग्णालय, नाशिक व अमरावती येथे दोन महिन्याकरीता डेमोकरीता बसविण्यांत यावे. तसेच त्याकरीता आवश्यक असलेले मनुष्यबळ आपल्या स्तरावरुन उपलब्ध करुन देण्यांत यावे.

(मा.आयुक्त महोदय यांची गान्य टिप्पणी)

प्रत - वैद्यकीय अधिक्षक, विभागीय संदर्भ सेवा रुग्णालय,नाशिक व अमरावती


२/- आपणांस कळविण्यांत येते की, उपरोक्त विषयी ई-मिङिका या उपकरणांचे प्रात्यक्षिक करण्याकरीता सदरहू मशीन आपल्या रुग्णालयामध्ये दोन महिन्याकरीता बसविण्यांत यावे. सदरहू मशीनचा उपयोगिता अहवाल, त्याची किंमत, तसेच रुग्णांना त्याचा होणार फायदा यावावतचा तज्ञांकडून अहवाल आयुक्तालयास सादर करण्यांत यावा.

प्रत -उपसंथालक. आरोग्य सेवा, नाशिक मंडळ, नाशिक / अकोला मंडळ, अकोला



CLINICAL TRIAL REPORT BY GOVERNMENT OF MAHARASHTRA













Maharashtra Government Health Department

Medical Superintendent, Regional Referral Services Hospital, (Daffrin Hospital Campus), Shrikrushana Peth, Amravati-444601

Ph.No. 0721-2970161, 0721-2970167. Outward no./RRSH,Amt./MOU /

3844 /2022

Date. 99/E/2022 Email: msrrshamt@gmail.com

To, Rohera Healthcare & Technologies Pvt Ltd E-Medica Pune.

In Regional Referral Services Hospital, Amravati the installation & inauguration of E-Medica device was done in the month of February 2022, since that time patients were enrolled of different diseases as per instructions given by experts of E-Medica.

Total of 132 Covid confirmed cases were enrolled & studied for their improvement in health during this study period. About 60 minutes sessions were given on respective described mode of E-Medica device & parameters like SPO2, Pulse, Temperature, Blood pressure were noted for three different settings.

A total 110 already diagnosed hypertensive patients & 153 diagnosed type II Diabetes mellitus patients were enrolled in study to see the effect of device to minimize blood pressure & blood sugar respectively.





Individual values of baseline (i.e. before use of device) & also values after the use of device on day one, day two & day three were noted in Microsoft excel. Mean and standard deviation were calculated & 't' test were applied using open epi software. After applying 't' test of significance, 'p' value less than 0.05 at 95% confidence interval was considered statistically significant.

From the study conducted in this hospital it was concluded that, use of E-Medica device in covid-19 positive patients is a potent adjuvant to reduce the body temperature, pulse rate & improve the oxygen saturation levels in very short continuance of its utility. Also E-Medica device can be adjuvant to decrease the blood pressure & Blood sugar level in diagnosed hypertensive & diabetic patients.

Above results are due to following reasons as stated in Information Brochure given by the manufacturer of E-Medica-

"After applying eMedica device on the specific program (as pre-programmed & fed in the device), with specified frequency, voltage & current, it will Enhance the Cell charge of the organ. This process Improves the function of the Organ, Reduces Blood-Fat, Reduces the chances of Cardiac Arrest, Improves the Blood Circulation. When a mild current is exposed to virus or bacteria it would destroy the Virus & Bacteria in the body. In diabetic patients eMedica device transmits specific Voltages and Frequency with the specific electric charge into the human body. As this 3-dimensional technology passed through the blood stream, it enhances the charge of the β-cell in the pancreas, which in-turn improves the function of the organ helping it to produce more insulin. As a specific combination of Voltages-Frequency-Current is passed through the blood stream, it burns the fat/deposit/plaque from the blood vessel. This allows the





insulin generated from the $(\beta$ -cell) pancreas to penetrate into the blood cells. As the natural insulin (produced by β -cell) reached the blood stream it immediately balances the Blood Sugar Level. Hence controlling Diabetes." (Reference-www.emedica.in. Robera Healthcare & Technology Private Limited ROHERA INC.)

This present pilot study was conducted on a limited sample size, which proved the efficacy on covid positive & hypertensive patients. Currently the study is being in process in large number of groups to see its results on Diabetic & hypertensive Patients. The long term effect of e- medica is conducted on Diabetic patients to understand better blood sugar control over a long term use. As soon we achieve our desired sample size, we will publish its results.

Detailed results of this study with mean with standard deviation & P value are enclosed with this.

Dr. M. R. Mendhe – OSD, RRSH, Amravati.

Dr. G. Khobragade – MO, RRSH, Amravati.

Dr. Ravi Bushan – Nodal Officer RRSH-Covid, Amravati


Dr. V. S. Pawar – MO, RRSH, Amravati.

5. Dr. V. S. Dhande- MO, TBH, Amravati.

Dr. Shyam Gawande- MO, RRSH, Amravati

Medical Superintendent RRSH, Amravati dical Superintendent

Medical Superintendent Regional Referral Services Hospital AMEAVATI.





COVID-19

Table No.1: Effect of Emedica device on temperature in covid 19 positive patients (n=132)

DAY	TEMPARATURE	P Value	
	BEFORE	AFTER	(t test)
DAY 1	99.99 ± 0.98	97.20 ± 0.88	<0.0000001
DAY 2	99.56 ± 0.64	97.25 ±0.94	<0.0000001
DAY 3	99.56 ± 0.51	97.61 ± 0.76	< 0.0000001


Table No.2: Effect of Emedica device on oxygen saturation in covid 19 positive patients

DAY	SPO2 (Mear	P Value	
	BEFORE	AFTER	(t test)
DAY 1	93.12 ± 1.52	96.15 ± 1.19	<0.000000
DAY 2	94.94 ± 2.68	97.06 ± 1.64	<0.0000001
DAY 3 94.78 ± 2.24		97.48 ± 1.30	< 0.0000001

Table No.3: Effect of Emedica device on pulse rate in covid 19 positive patients

DAY	PULSE (Mea	P Value	
	BEFORE	AFTER	(t test)
DAY 1	92.09 ± 5.76	84.28 ± 12.51	<0.0000001
DAY 2	94.68 ± 4.91	89.27 ± 7.42	<0.0000001
DAY 3	96.12 ± 4.48	90.53 ± 5.73	<0.0000001

Total 132 covid positive patients were enrolled in research to study the effect of Emedica device on improvement in health status of patients. It was found that mean temperature of covid positive patients get decreases, after a session of 60 minutes duration in a day, also it was found that mean temperature of patients were decreases subsequently on day 2 & 3 after its use and the difference was statistically significant. (P <0.05).





Similarly, one of the crucial parameter in covid patients i.e. oxygen saturation shows drastic increase in SPO2 & the difference was statistically significant. (P<0.05)

Mean pulse was also reduced subsequently after the use of Emedica device, & difference was also statistically significant.


Thus, it can be concluded that, use of Emedica devise in covid 19 positive patients can be a potent adjuvant to reduce the body temperature, pulse rate & improve the oxygen saturation in very short continuance of its utility.

RANDOM BLOOD SUGAR

Table No.6: Effect of Emedica device on random blood sugar in diagnosed case of diabetes mellitus. (n=153)

DAY	RANDOM BLOOD SU	GAR (Mean with SD)	P Value	
	BEFORE	AFTER	(t test)	
DAY I	197.16±20.00	136.84 ± 16.92	<0.0000001	
DAY 2	185.19 ± 19.75	135.17 ± 16.23	<0.0000001	
DAY 3	172.90 ± 21.55	126.56 ± 15.58	<0.0000001	

Total 153 diagnosed cases of Diabetes mellitus type 2 were studied for the effect of Emedica device on blood sugar level & it was found that after a session of Emedica device random blood glucose level get decreases and the difference was statistically significant. (P < 0.05)





BLOOD PRESSURE

Table No.4: Effect of Emedica device on systolic blood pressure in diagnosed case of hypertension (n=110)

DAY		OOD PRESSURE with SD)	P Value (t test)
	BEFORE	AFTER	
DAY 1	147.73 ± 15.63	125.93 ± 16.45	< 0.0000001
DAY 2	139.70 ± 11.55	131.5 ± 13.52	0.000002496
DAY 3	143.30 ± 9.12	133.94 ± 9.73	< 0.0000001

Table No.5: Effect of Emedica device on diastolic blood pressure in diagnosed case of hypertension

DAY	OIASTOLIC BLOOD PRESSURE (Mean with SD)		P Value (t test)
	BEFORE	AFTER	
DAY I	90.21 ± 9.87	88.85 ± 13.70	0.3992
DAY 2	84.71 ± 8.00	74.77 ± 6.89	< 0.0000001
DAY 3	86.72 ± 7.86	82.68 ± 7.70	0.0001547


Total 110 diagnosed cases of Hypertension were studied to see the effect of Emedica device on its effect on lowering blood pressure. It was found that both systolic and diastolic blood pressure get decreases after the use of Emedica device. Also the difference was statistically significant. (P < 0.05)



CALIBRATION REPORT







info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTC

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com

Phone: 02138 - 232901, 232902, 232903 CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2: +91 7768005411

Mobile 3: +91 7768005422

TEST REPORT

Page No. 1 of 10

TEST REPORT AS PER :- IS 13450 (Part 1)/ IEC 60601-1

SRF No.: 21040014

e-Medica Office No A3/A4/A5 Royal Arcade, Pune Satara Road, Near Shankar Maharaj Math, Pune - 411043

Name & Address of Customer:

ULR No.: TC510021000000450F Discipline: Electrical Group; Safety Testing Facility Test Report No: HPLI/Test/2104001401 Date of Issue: 04/05/2021

Customer Ref. & Date: 09/04/2021

Date of Sample Receipt:

09/04/2021

Start of Test Date: 09/84/2021

N/A

N/A

WE PRESENT APPROATES

PUCTA PL

Approved by

End of Test Date: 30/04/2021

PART A -PARTICULARS OF THE SAMPLE SUBMITTED

Sample description	e-Medica (Wellness Device)
Grade/ variety/ type/ class/ size etc.	e-Medica (Wellness Device)
Declared values, if any	e-Medica (Wellness Device):- Input Rating: 19.0Vdc, 1.0A
	AC/DC Adapter:- Input Rating: 100-240Vac, 50/60Hz,0.3A
	Output Rating: 19.0Vdc,1.0A
Code no., BIS seal and IO's sign, if any	NII
Batch no., date of manufacture and	Model No. e-Medica (Wellness Device)
Brand name	Brand Name: e-Medica
Quantity	01
Condition of the sample	Good
Reference specification (s)	IS 13450 (Part 1)/ IEC 60601-1
ES) INVIN	(Tests have been carried out as per customer request)
Environmental conditions	Temperature (25±5)°C & Relative humidity <70%

PART B - SUPPLEMENTARY INFORMATION

Statement of Conformity

Decision Rule

- a) Deviations from the test methods as per relevant specification/ work instructions, if any:
- b) Details of the drawings, graphs, tables, sketches or photographs as referred in the test report, if any: ANNEXURE A attached
- c) Testing procedure according to work instruction : HPLI03/Test-Medical/WI-01 to 07
- d) The Management System is maintained in accordance with IS/ISO/IEC 17025:2017 and testing Standards/Instruments are traceable to National/International Standards.
- Notes: i) This report is not to be reproduced wholly or in part without our special permission in writing.
 - ii) This report refers only to the particular sample detailed above.
 - iii) The results reported in this certificate are valid at the time of and under the stipulated conditions of measurement.

Checked by

Format No - HPLL04 F 11-

Registered Office: B-9/51, Sector-18, Rohini, Delhi - 110089







B-32/1/2, MIDC, Ranjangoon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1 : +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ULR No.: TC510021000000450F Test Report No.: HPLI/Test/2104001401

PART C

TEST RESULT

Sr. No.	TESTS WITH CLAUSE REFERENCE	SPEC	TIFIED REQUIREMENTS	IS 13450 (Part 1)/IEC 60601 RESULTS
1.	Power Input Test (Cl.4.11 of IS 13450 (Part 1) /IEC 60601-1)	i)	Steady-state measured input of ME Equipment or ME System at rated voltage and at operating settings indicated in instructions for use did not exceed marked rating by more than 10%	See appended Table 4.1]
		ii)	-Measurements on ME Equipment or ME System marked with one or more rated voltage ranges made at both upper and lower limits of the range	See appended Table 4.11
2.	Humidity Preconditioning Treatment (CL5.7 of IS 13450 (Part 1) /IEC 50601-1)		ME Equipment or parts there of affected by climatic conditions were set up completely, or partially, with covers detached and subjected to a humidity preconditioning prior to tests of Clauses 8.7.4 and 8.8.3.	Before the test of 8.7.4 & 8.8.3, the humidity preconditioning is performance
			ME Equipment heated to a temperature between T and T + 4 °C for at least 4 h and placed in a humidity chamber with a relative humidity of 93 % ± 3 % and an ambient within 2 °C of T in the range of + 20 °C to + 32 °C for 48 h	25 degree C 93 %RH
3.	Patient Auxiliary Currents (CL8.7 of IS 13450 (Part 1) /IEC 60601-1)	8.7.1	a) Electrical isolation providing protection against electric shock limits currents to values in 8,7,3	See appended Table 8.7
			b) Specified values of Earth leakage, Touch, Patient Leakage, and Patient Auxiliary Currents applied in combination of conditions in appended Table 8.7	See appended Table 8.7
		8.7.2	Allowable values specified in 8.7.3 applied under Single Fault Conditions of 8.1 b), except	See appended Table 8.7
		Where insulation used in conjunction with a Protective Earth Connection, insulation short circuited only under conditions in 8.6.4 b). The only single Fault Condition for Earth Leakage current was interruption of one supply conductor at a time. Leakage currents and patient auxiliary current not measured in single fault condition of short circuiting of one constituent part of double insulation.	See appended Table 8.7	
			Leakage current was interruption of one supply	See appended Table 8.7
			auxiliary current not measured in single fault condition of short circuiting of one constituent	See appended Table 8.7
			Single fault conditions not applied at same time as special test conditions of maximum mains voltage on applied parts and non-projectively earthed parts of enclosure	See appended Table 8.7

Checked by/ Page 02 of (0 HI PHYSIX LABORATORY INDIA PUT. LTD.

Approved by



info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTC

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, Infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1 : +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ULR No.: TC510021000000450F Test Report No.: HPLI/Test/2104001401

Sr.		SPEC	TIFIED REQUIREMENTS	IS 13450 (Part 1)/IEC 6060 RESULTS
	. NEW ENCINCE	8.7.3	 a) Allowable values in 8,7.1 b), c), and d) measured based on, and are relative to currents in Fig 12 a), or by a device measuring frequency contents of currents as in Fig 12 b; 	See appended Table 8.7
			 b) allowable values of patient leakage and auxiliary currents are according to tables 3 & 4, and values of a.c. are relative to currents having a frequency not less than 0.1Hz; 	See appended Table 8.7
			 c) Touch current did not exceed 100 μA in normal condition and 500 μA in single fault condition 	See appended Table 8.7
			 d) Earth Leakage Current did not exceed 5 mA in normal condition and 10 mA in single fault condition 	See appended Table 8.7
			Higher values of Earth Leakage Current permitted for Permanently Installed ME Equipment connected to a supply circuit supplying only this ME Equipment according to local regulations or IEC 60364-7-710:	See appended Table 8.7
			 c) Leakage Currents, regardless of waveform and frequency, did not exceed 10 mA r.m.s. in normal or in single fault condition (measured with a non- frequency-weighted device) 	See appended Table 8.7
		8.7.4	Leakage and Patient Auxiliary Currents measurements:	See appended Table 8.7
4.	Temperature Rise Test (Excessive temperatures in ME Equipment) (Cl.11,1 of IS 13450 (Part 1) /IEC 60501-1)	11.3.1	Temperatures on ME Equipment parts did not exceed values in Tables 22 and 23 operating in worst-case normal use at maximum rafed ambient operating temperature T	See appended Table 11.1.1
			Surfaces of test corner did not exceed 90 °C	Satisfactory (See appended Table 11.1.1)
		11.1.2	Applied pans (hot or cold intended to supply heat to a patient comply	Applied part not supply heat to patient
		11.1.2.2	patient complies with the limits of table 24 in both normal condition and single fault condition	Complies (See appended Table 11.1.1)
			Applied parts surface temperature exceeds 41°C disclosed in the instruction manual:	Not exceed 41°C
			Applied parts surface temperature of equal to or less than 41°C	See appended Table 11.1.1
· /	Interruption of Power Supply (Cl.11.8 of IS 13450 (Part I) /IEC 60601-1)		Interruption and restoration of power supply did not result in a hazardous situation, except interruption of its intended function	Unit operation normally after Interrupted and restored (Equipment state: Loaded, operating mode)

Wested by

Page 03 of 10

Checked by



™ info.emedica@gmail.com



HI PHYSIX LABORATORY INDIA PVT. LTC

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra into@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ULR No.: TC510021000000450F Test Report No.: HPLI/Test/2104001401

Sr. No.		SPECI	FIED REQUIREMENTS	IS 13450 (Part 1)/IEC 60601 RESULTS
6.	Abnormal operation & Single Fault Condition (Cl.13 of IS 13450 (Part 1) /IEC 60601-1)		None of hazardous situations in 13.1.2-13.1.4, inclusive, occurred when single fault conditions applied, one at a time, as in 4.7 and 13.2	Satisfactory
		13.1.2	Emission of flames, molten metal, poisonous or ignitable substance in hazardous quantities did not occur	No fire, emission of molten metal or ignition of substances was noted during the tests.
			Temperatures of applied parts did not exceed allowable values in table 24 when measured as in 11.1.3	Complies (See appended Table 11.1.1)
			Temperatures of ME equipment parts that are not applied parts likely to be touched did not exceed values in table 23 when measured and adjusted as in 11.1.3	See appended Table 11.1.1
			Limits for leakage current in single fault condition based on 8.7.3 did not exceed	Complies (See appended Table 8.7)
			Voltage limits for accessible parts including applied parts in 8.4.2 did not exceed	Complies (See appended Table 8.7)
		13,2.1	During application of single fault conditions in 13.2.2-13.2.13, inclusive, normal conditions in 8.1 a) applied in least favourable combination	Considered
		13.2.2 to 13.2.12	ME Equipment complied with 13.2.2 -13.2.12	See appended Table 13.2
		13.2.13	ME Equipment remained safe after tests of 13.2.13.2 to 13.2.13.4 (inclusive), and cooling down to within 3°C of the temperature in the test environment	See appended Table 13.2
			For insulation of thermoplastic materials relied upon as a means of protection (see 8.8), the ball pressure test specified in 8.8.4.1 a) performed at a temperature 25 °C higher than temperature of insulation measured during tests of 13.2.13.2 to 13.2.13.4 (inclusive).	Bull-pressure test for nonmetallic enclosure has been carried out at 75°C & found Satisfactory (See appended Table 8.8.4.1)
7.	Drop test (CL15.3.4 of IS 13450 (Part 1) /IEC 60601-1)	15.3.4.2	Sample of portable ME Equipment, accessories and portable part with safe working load lifted to a height as in table 29 above a 50 ± 5 mm thick hardwood board lying flat on a concrete floor or rigid base, dropped 3 times from each orientation in normal use (cm)	There was no damage to the interior or exterior of the tested sample which caused exposure to live parts or increased risk of electric shock

PART D:

Remarks:

1. The observations given in part A of the cover page of the test report are taken from the marking on the sample and specification given by the customer.

2. All the test covered in this test reports are performed as per Customer request.

In Physix Lisoratory INDIA PUT. LTD.

flite Tested by

Checked by

Page 04 of 10/

Approved by L'ectronics division)



info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTC

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra Info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ANNEXURE A

ULR No.: TC510021000000450F Test Report No.: HPL1/Test/2164001401

8.10	TABLE: List Of Critical Components							
Object/part or Description	Manufacturer/ Trademark	Type/Model	Technical data	Standard (Edition/year)	Mark(s) of Conformity ¹)			
AC/ DC Adapter	Magnetics	LT19-SEM-7450	Input :100-240Vac .50/60Hz,0.3A Output: 19Vdc, 1.0A	IS13450 (Part 1)/ IEC60601-1	Tested with appliance			
Interconnecting Cable	R R Kabel Limited	Various.	PVC 1100V	IS 694	CN/L. 3934068			
LCD Display	Shenzhen Jing Handa Electronics. Ltd	JHD204A	5.0V, 1.2mA	IS13450(Part I)/ IEC60601-1	Tested with appliance			
Varistor	TDK Electronics GmbH & Co OG	S14 series	105°C, up to 8kV 30Vrms,38Vdc	IEC 61051-1 IEC 61051-2 IEC 61051-2-2	VDE 40027582			
Plastic Enclosure	Various formation: 11 An aste	Various	Plastic, V-1, min. 1.5 thickness, 80°C	UL94	UL			

Licenses and Certificates of Conformity for verification.

4.11 TABLE: Power Input					
Operating Conditions / Ratings	Voltage (V)	Frequen cy (Hz)	Current (A)	Power (W or VA)	Power factor (cos φ)
Test on Model e-Medica	-10				47
For AC/DC Adapter: (Shyam Electronics Magnetics Type LT19-SEM-7450)			T	-	-
Max. normal loaded / 1.0 A	90Vac	50	0.318	22.59	
Max. normal loaded / 1.0 A	90Vac	60	0.320	22.60	
Max. normal leaded / 10 A	100Vac	50	0.306	22.45	_
Max. normal loaded / 1.0 A	100Vac	60	0.309	22.48	/
Max. normal loaded / 1.0 A	240Vac	50	0.196	22.38	
Max. normal loaded /1.0 A	240Vac	60	0.199	22.40	
Max. normal loaded / 1.0 A	264Vac	50	0.194	22.34	74
Max. normal loaded /1.0 A	264Vac	60	0.197	22.35	
Max. normal loaded	19Vdc		0.556	10.56	-

Lete by

Page 05 of 10



info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTC

8-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com

Phone: 02138 - 232901, 232902, 232903 CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2: +91 7768005411

Mobile 2: +91 7768005411 Mobile 3: +91 7768005422

ULR No.: TC5100210000000450F Test Report No.: HPLI/Test/2104001401 IS 13450 (Part 1)/IEC 60601-1

8.7 TABLE: Leakag		7	-	
Type of leakage current and tes condition (including single faul	t Supply voltage (V)	Supply frequency (Hz)	Measured max.value (µA)	Remarks
Fig. 13 - Earth Leakage (ER)	-			Maximum allowed values: 5 mA NC; 10 mA SFC
Before Humidity Condition:				The second secon
ER, NC	264	60	152.3	S1 = 1, S5 = 1
ER, NC	264	60	157.9	S1 = 1, S5 = 0
ER, SFC (Neutral Open),	264	60	197.5	S1 - 0, S5 = 1
ER, SFC (Neutral Open)	264	60	199.1	S1 - 0, S5 = 0
After Humidity Condition:			(SCOM)	3. V, U3 - U
ER, NC	264	60	153.8	S1 = 1, S5 = 1
ER, NC	264	60	157.7	S1 = 1, S5 = 1 S1 = 1, S5 = 0
ER, SFC (Neutral Open),	264	60	198.4	
ER, SFC (Neutral Open)	264	60	199.5	S1 = 0, S5 - 1
(with non frequency-weighted				SI = 0, S5 - 0
Before Humidity Condition:				Maximum allowed values: 10 mA for NC & SFG
ER. NC	264	60	160.1	
ER, NC	264	60	160.1	S1 = 1, S5 = 1
ER, SFC (Neutral Open),	264	60	201.6	S1 = 1, S5 = 0 S1 = 0, S5 = 1
ER, SFC (Neutral Open)	264	60	205.7	S1 = 0, S5 = 1 S1 = 0, S5 = 0
After Humidity Condition:			20011	131-4,33-4
ER, NC	264	60	160.8	SI = 1, S5 = 1
ER, NC	264	60	164.2	SI = 1, S5 = 0
ER, SFC (Neutral Open),	264	60	201.5	S1 = 0, S5 = 1
ER, SFC (Neutral Open)	264	60	205.4	S1 = 0, S5 = 0
ig. 14 - Touch Current (TC)				Maximum allowed values: 100 μA NC; 500 μA
MD between Earth and AC/DC A	dapter Piast	ice Enclosur	e surface (co	vered with metal foil)
Before Humidity Condition:				/
C, NC	264	60	7.6	\$1 = 1, \$5 = 1, \$7 = 1
C, NC	264	60	5.8	S1 = 1, S5 = 0, S7 - 1
C, SFC (Neutral Open)	264		9.1	S1 = 0, S5 = 1, S7 = 1
C, SFC (Neutral Open)	264		9.5	S1 = 0, S5 = 0, S7 - 1
C, SFC (Ground Open)	264	60	8.3	\$1 - 1, \$5 = 1, \$7 = 0
C, SFC (Ground Open)	264	60	17.7	S1 = 1, S5 = 0, S7 = 0

Wested by

Page 06 of 10

Checked by



info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTC

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754 Mobile 1: +91 7768005400

Mobile 2: +91 7768005411 Mobile 3: +91 7768005422

ULR No.: TC510021000000450F Test Report No.: HPLI/Test/2104001401

Type Of Leakage Current And Test Condition (Including Single Faults)	Supply Voltage (V)	Supply Frequency (Hz)	Measured Max. Value (Ma)	Remarks
After Humidity Condition:				
TC, NC	264	60	8.3	S1 = 1, S5 = 1, S7 = 1
TC, NC	264	60	6.1	SI = 1, S5 = 0, S7 = 1
TC, SFC (Neutral Open)	264	60	9.5	S1 = 0, S5 = 1, S7 - 1
TC, SFC (Neutral Open)	264	60	9.8	S1 = 0, S5 - 0, S7 = 1
TC, SFC (Ground Open)	264	60	8.8	S1 = 1, S5 = 1, S7 = 0
TC, SFC (Ground Open)	264	60	19.1	S1 = 1, S5 = 0, S7 = 0
Fig. 15 - Patient Leakage Current (P) MD between surface of footprint o earth)	•		-	Maximum allowed values: Type B or BF AP: 10 μΑ NC; 50 μΑ SFC (d.c. current); 100 μΑ NC; 500 μΑ SFC (a.c.) Type CF AP: 10 μΑ NC; 50 μΑ SFC (d.c. or a.c. current)
P, NC	264	60	ac<1	S1 = 1, S5 = N
P, NC	264	60	ac<1	S1 = 1, S5 = R
P, SFC	264	60		S1 = 0, S5 = N
P,SFC	264	60	ar: 1.32 nA	\$1 = 0 \$5 = B
Fig. 16 - Patient leakage current wi PM) (MD between External voltage	th mains o ge & surfa	n the Fatone	applied parts of)	Maximum allowed values: Type B: N/A , Type BF AP: 5000 μA
P. NC	264	60		Type CF AP: 50 µA S1 = 1, S5 - N
, NC		100		S1 = 1, S5 = N S1 = 1, S5 = R
P, SFC	PASSAGE	No.	the same of the sa	S1 = 0, S5 = N
P.SFC	and the same of	100		S1 = 0, S5 = R

upplementary information:

Note 1: For Earth Leakage Current see 8.7.3 d) and 8.7.4.5;

Note 2: For Touch Current sec 8.7.3 c) and 8.7.4.6;

Note 3: For Patient Leakage Current See 8.7.3.b) and 8.7.4.7

Note 4: Total Patient Leakage Current values are only relative to equipment with multiple Applied Parts of the same type. See 8.7.4.7 h). The individual applied parts complied with the Patient Leakage Current values.

Note 5: In addition to conditions indicated in the Table, tests conducted at operating temperature and after humidity preconditioning of 5.7, equipment energized in stand-by condition and fully operating, max rated supply frequency, at 110 % of the max rated mains voltage

ER - Earth leakage current TC - Touch current

P - Patient leakage current

PA - Patient auxiliary current

TP - Total Patient current

PM - Patient leakage current with mains on the applied parts

MD - Measuring device

A - After humidity conditioning

B - Before humidity conditioning

1 - Switch closed or set to normal polarity

0 - Switch open or set to reversed polarity

NC - Normal condition SFC - Single fault condition



Page 07 of 10





™ info.emedica@gmail.com



HI PHYSIX LABORATORY INDIA PVT. LTE

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1 : +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ULR No.: TC5100210000006450F Test Report No.: HPLI/Test/2104001401

IS 13450 (Part 1)/IEC 50601-1

8.8,4.1	TABLE: Resistance to heat - Ball pressure test of thermoplastic parts							
	41147		2 mm	+				
	Force (N):	20			-			
Part/material Plastic Enclosure			Test temperature (°C)	Impress	ion diameter (mm)			
			75	0.8				
Supplemen	ntary information:			1	777-17-17-17			
The results	were less 2 mm. Cracks were visible to the naked	Love						

11.1,1	TABLE: Excessive temperatures in ME Equipment						
Model N		See below	4+			**	
Test ami	bient (°C);	25.6	-	-		_	-
Test sup	ply voltage/frequency (V/Hz)(4);	90V/50Hz	94	4		12	
Sr.No	Thermccouple location(3)	le location(3)		Max allowable temperature ¹⁷ from Table 22, 23 or 24 or RM file for AP(5) (°C)		measured ature(2),	Remarks
1	PCB		105		31.4		35.8
2	Plastics Enclosure		60		3	0.9	35.3
3	LCD Display		60		2	7.9	32.3
4	LCD Display PCB		105		2	7.4	31.8
5	Applied part		41 26.8		5.8	31.2	
6	Adapter Enclosure		60 35.8		5.8	40.2	
7	Internal wire		105 28		8.7	33.1	
8	Metal		56 31.3			35.7	
9	Test Corner		90 25.9			30,3	
10	Ambient				100	5.6	Shift Tma 30

Supplementary information: ¹³When thermocouples used to determine temperature of windings, limits of Table 22 reduced by 10 °C.

Tested by

Page 08 of 10

Checked by /



™ info.emedica@gmail.com





HI PHYSIX LABORATORY INDIA PVT. LTD

B-32/1/2, MIDC, Ranjangoon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2: +91 7768005411 Mobile 3: +91 7768005422

ULR No.: TC510021000000450F Test Report No.: HPLL/Test/2104001401

IS 13450 (Part I)/IEC 60601-1

13.2	TABLE: Single Fault Conditions		
Clause No.	Description of Single Fault Condition	Results observed	Hazardous Situation (Yes/No)
13.2.2	Electrical Single Fault Conditions per Clause 8,1:	-	*
	Output shorted (AC/DC Adapter)	Unit switched off. No hazard, No defect.	No
	Output overload	Output current: 1.48A max. Maximum temperature of the enclosure: 46.8°C at 25.5°C ambient temperature.	No
		No hazard. No defect.	

Supplementary information:

The measured temperatures did not exceed those allowable. See Table 11 for Temperatures obtained during the indicated Abnormal Operation tests.

There was no flame, extensive smoke or melted metal.

Page 09 of 10



info.emedica@gmail.com
 info.emedica@gmail.com



HI PHYSIX LABORATORY INDIA PVT. LTD

B-32/1/2, MIDC, Ranjangaon, Pune, Maharashtra info@hiphysix.com, infohplindia@gmail.com Phone: 02138 - 232901, 232902, 232903

CIN: U74120DL2009PTC194754

Mobile 1: +91 7768005400 Mobile 2 : +91 7768005411 Mobile 3 : +91 7768005422

ULR No.: TC510021000000450F Test Report No.:HPLI/Test/2104001401

IS 13450 (Part 1)/IEC 60601-1





External View



Internal View



AC/DC Adapter

AC-DC POWER ADAPTER

Model 1719 SEM 7450 MPUT: 100VINC 240VAC 50/60Hz 0.3A OUTPUT: 19 - 1A сытон for our with Industrial Technology Equipment
ATTENTION
VIOLES AND DU MATERICAN PRIMATIONS BUILDINGS PARA LISE CONTROL POS DE VECHOLOGIA DE LA INFORM

Copy of Marking Plate of AC/DC Adapter



Trademark

Checked by

Page No. 10 of 10


***** END OF THE TEST REPORT *****

WE PHYSER LABORATORY INDIA PVT. LTD. Approved by



CERTIFICATES











National Accreditation Board for Testing and Calibration Laboratories

NABL

CERTIFICATE OF ACCREDITATION

HI PHYSIX LABORATORY INDIA PRIVATE LIMITED

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

B-32/1/2, MIDC, INDUSTRIAL AREA, RANJANGAON, PUNE, MAHARASHTRA, INDIA

in the field of

TESTING

Certificate Number:

TC-5100

Issue Date:

19/11/2021

Valid Until:

18/11/2023

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Entity: HI PHYSIX LABORATORY INDIA PRIVATE LIMITED

Signed for and on behalf of NABL



N. Venkateswaran Chief Executive Officer





Certificate of Compliance

RoHS Directive (2011/65/EU) of the European Parliament and of the Council on the restriction of use of certain Hazardous Substances in Electrical and Electronic Equipments

Certificate No. QVA-RLHV-22-1517483

Name : ROHERA HEALTHCARE AND TECHNOLOGY PVT LTD

Address : FLAMINGO, 101 RAHEJA GARDENS, WANORIE PUNE-411040, INDIA.

Product : "MANUFACTURER OF ELECTRONIC MEDICAL DEVICE E-MEDICA"

This is to state that the above mentioned products is in compliance with RoHS Directive (20/95/EC) of the European Parliament and Commission Decision 2005/618/EC on the restriction of use of certain Hazardous Substances [Lead (Pb), Mercury (Hg), Cadmium (Cd), Hexavalent Chromium (Cr6+), Polybrominated biphenyls (PBBs) and Polybrominated Diphenyl ethers (PBDEs)] in Electrical and Electronic Equipments.

Statement:

This certificate declares that the product type/model described above complies with the mentioned above European Standard(s).

Remarks

This certificate of complies is based on the evaluation of a sample of the above mentioned products. It does not imply and assessment of the mass-production of the product. This certificate holder may use this certificate in connection with the test certification body should be informed (revision of technical file) for any modification or alterations made to the aforementioned product type(s), including design and manufacture and/or extension to the existing scope of application.

The certificate is valid for three years if the company applies the technical construction file which has been stored QVA office. This certificate includes declaration of manufacturer.

Certificate remains property of QVA to whom it must be returned upon request. The certificate validity is conditioned by positive results or surveillance audits.

Validity of this certificate can be verified at www.gaafs.us

Date of Certification 15TH DEC 2022

1st Surveillance Due 14TH DEC 2023

2nd Surveillance Due 14TH DEC 2024

Certificate Expiry (Subject to the company maintaining its system 14TH DEC 2025

To the required standard)









CAB Address: Maryland Avenue, SW Washington, D.C. 20202

Validity of this certificate is subject to annual surveillance audits to be done successfully This certificate is the property of QVA Certification and shall be returned immediately on request

QVA Certification is an independent Systems Products and Personal assessment Body, QVA Certification is a accredited by GAAFS.US



info.emedica@gmail.com



We hereby declare that the technical file of product complied with the requirement of directives Medical Devices Directive 93/42/EEC

Applicants

Name : ROHERA HEALTHCARE AND TECHNOLOGY PVT LTD

Address : FLAMINGO 101 RAHEJA GARDENS WANORIE PUNE 411040

Product ___ : "MANUFACTURER OF ELECTRONIC MEDICAL DEVICE E-MEDICA"

Complies with the requirements applicable to it

The Certification body has performed an audit of the above product quality system covering the design, manufacture and final inspection of the certified product. The quality system has been assessed, approved and is subject to continuous surveillance according to the Medical Devices Directive 93/42/EEC

This certificate is issued under the following conditions:

It applies only to the quality system maintained in the manufacture of above referenced models and it does not substitute the design or type-examination procedures, if requested.

- The certificate remains valid until the manufacturing conditions or the quality systems are changed.
- 2. The certificate validity is conditioned by positive results or surveillance audits.
- After fulfilling the relevant EU legislation, the manufacturer shall affix to each device, of the above referenced models.
- 4. The CE mark as shown above can be used, under the responsibility of the manufacturer, after completion of an EC Declaration of conformity and compliance with all relevant EC Directives. The statement is based on a single evaluation of one sample of above mentioned product. It does not imply an assessment of the whole production.

Certificate No. QVA-RLHV-22-1517481

Certificate can be verified at www.gaafs.us

Date of Certification	15 TH DEC 2022
1 st Surveillance Due	14TH DEC 2023
2 nd Surveillance Due	14TH DEC 2024
Certificate Expiry (Subject to the company maintaining its system	14TH DEC 2025

To the required standard)











This is to certify that

ROHERA HEALTHCARE AND TECHNOLOGY PVT LTD

FLAMINGO, 101 RAHEJA GARDENS, WANORIE PUNE-411040, INDIA.

has been independently assessed by QVA and is compliance with the requirement of the standard

ISO 13485:2016

Quality Management System For Medical DevicesFor the following scope of activities

"MANUFACTURER OF ELECTRONIC MEDICAL DEVICE E-MEDICA"

Certificate Number: QVA-RLHV-22-1517482

To verify this certificate please visit at www.gaafs.us

Date of Certification 15TH December 2022 Issuance Date 15TH December 2022 1st Surveillance Due 14TH December 2023 2nd Surveillance Due 14TH December 2024 Re-Certificate Due 14TH December 2025












QVA Certification is an independent Systems Products and Person assessment Body, QVA Certification is a accredited by GAAFS.US



RECOGNITION





■ info.emedica@gmail.com





info.emedica@gmail.com
 info.emedica@gmail.com



27th September, 2022

Letter of Recommendation

To, eMedica 318 Morya Estate Veera desai Road, Andheri west, Mumbai 400102

We are glad to inform that the eMedica Wellness Device used at Department of Physiotherapy, Sports Medicine and Rehabilitation, nanavati Max Super Specialty Hospital for the last 8 months. The device has been used as a part of holistic treatment approach to patients suffering from Chronic pain and Neurological conditions.

The following effects were seen:

- Improvement in Sleep Patterns
- Reduction in muscle cramps

It is an innovative therapeutic device which can be used safely as a part of providing holistic treatment to patients for promotion and improvement of overall quality of



Regards,

Dr Ali Irani,

HOD, Physiotherapy, Rehabilitation & Sports Medicine, Nanavati Super Specialty Hospital

Head of Department :

Physiotherapy & Sports Medicine.



info.emedica@gmail.com

LNM Group Lions Hospital & Research Centre - Bhuj

For the People by the People

To The Team

eMedica Rohera Healthcare & Technology Private Limited

It is with immense pleasure I would like to share my feedback while using your device eMedica in my hospital. I appreciate the effort made by your team member Mr. Bhushan Joshi and Mrs. Bhakti Joshi from Kedar Global, who has explained the science & given us the device to use on various patients.

We are glad to share that after using the device on various patients we have found promising & positive results. We conducted 30 days trials in our hospital with Mr. Sitaram Salpekar & Mr. Dashrath Kumar, They are the dialysis Technicianof LNM Lions Hospital and Research Center, Also few patients were recommended by Lions hospital team who received the treatment and saw the results. We conducted the treatment on them with eMedica Device as per their health conditions.

Mr. Sitaram Salpekar, 51 years old, a Dialysis Technicianwas quite low on energy; could not get proper sleep at night and did not have urge to eat.

As prescribed by eMedica Advisor he was given the Immunity program; Detox of Kidney &LiverProgram. Virus Infection Program, and finally the Cold & CoughProgram. The result we witnessed in 30 days was quite appreciated by the patient & our doctors. Hewas feeling Energetic, was able to walk for a longerperiod; Started getting better sleep; Felt hungrier; and water thirst increased. Trial date is from 22th June 2022 to 22th July

Mr. Dashrath Kumar, 27 years old, a Dialysis Technician was not well and on testing we found his platelets were low.

So, to begin with we treated him with eMedica and given the Immunity Program and Virus Program. To our surprise his platelets went up from 172000 to 210000 plus and his energy levels were very high. Trial date is from 10th July 2022 to 22th July 2022.

Ms. Hiral Thakkar, 22 Years old, a dialysis Patient suffering from Hypertension. After using eMedica she felt better in terms of energy and anxiety. Also, the fluctuation in Hypertension has reduced. Trial date is from 11th July 2022 to 20th July 2022.

Mr. Paresh Joshi, 47 years old, suffering fromParathyroid. He used eMedica twotimes a day-Immunity Program & Thyroid Program. Felt much energized & helped in getting better sleep. Thyroid tests yet to be done. Trial date is on 4th July to 8° July 2022.

Mrs. ManglabenMakhwana, 50 years old. Suffering from diabetes. Mrs. Vanitaben Savji Patel above 55 years old, suffering from sciatica,

Mrs. DhanbaiBhimjiKerai 69 years old a Dialysis Patient, suffering from Diabetes, Hypertension, Thyroid, Digestive Gas issue. Trials are on with them. Yet to discover results.

The patients are highly satisfied with eMedica usage and they would like to pursue the treatment further.

Vyoma Mehta LNM GROUP LIONS' HOSPITAL AND RESEARCH CENTER - BHUJ



■ info.emedica@gmail.com



A Unit of

TO WHOM SO EVER IT MAY CONCERN

I Dr. Mohan Kumar V have been introduced to eMedica and found the concept quite interesting and since it has all the certification in place I decided to use eMedica on my patients.

On 10th of April 2022 I started using eMedica on my patient Mrs.Saraswathi Her right side of the body is paralyzed. She was unable to move her right side at all.

To test the device (after making sure the device has no side effect). I made her use the device regularly for the paralytic program daily 4 hours from 10th April 2022 to 30th of April 2022.

After using for only 20 days, to my surprise I found excellent results. She can move & lift her right hand. I suggest her to continue the same for few more months to get back to normal.

I highly recommend the device for Paralysis.


The Revive
A Hollstic Neuropsychiatric Care
(A Unit of CMD Hospital)
9, Beside Hebbal Lake, Near V-Leed,
Hebbal Ring Road, Mysore-570016

□ 0821-2955188 | 81238 29876
 □ therevivehospital@gmail.com
 □ therevive in

W9, Seside Hebbal Lake,
 Near V-LEAD Hebbal Ring Road
 Mysuru - 570016



PATENT





@ www.emedica.in

info.emedica@gmail.com

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

WIPO PCT

(19) World Intellectual Property Organization

International Bureau

(43) International Publication Date 18 March 2021 (18.03.2021)



(10) International Publication Number WO 2021/048642 A1

(51) International Patent Classification:

A61B 5/00 (2006.01) G16H 50/20 (2018,01) G16H 20/00 (2018.01)

A61N 1/36 (2006.01) A61N 2/00 (2006.01) A61B 17/00 (2006.01)

G16H 10/60 (2018.01)

(21) International Application Number:

PCT/IB2020/055292

(22) International Filing Date:

05 June 2020 (05.06.2020)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

201921036412

10 September 2019 (10,09,2019) IN

- (72) Inventor; and
- (71) Applicant: ROHERA, Hemant Karamchand [IN/IN]: 101, Flemingo Raheja Gardens, Wanowrie, Pune-Maharashtra 411040 (IN).
- (74) Agent: DEWAN, Mohan; R. K. Dewan & Co. Podar Chambers, S. A. Brelvi Road, Fort, Mumbai-, Maharashtra 400001 (IN).
- (81) Designated States (unless otherwise indicated, for every kind of national protection available): AE, AG, AL, AM, AO, AT, AU, AZ, BA, BB, BG, BH, BN, BR, BW, BY, BZ, CA, CH, CL, CN, CO, CR, CU, CZ, DE, DJ, DK, DM, DO, DZ, EC, EE, EG, ES, FI, GB, GD, GE, GH, GM, GT, HN, HR, HU, ID, IL, IN, IR, IS, JO, JP, KE, KG, KH, KN, KP, KR, KW, KZ, LA, LC, LK, LR, LS, LU, LY, MA, MD, ME, MG, MK, MN, MW, MX, MY, MZ, NA, NG, NI, NO, NZ, OM, PA, PE, PG, PH, PL, PT, QA, RO, RS, RU, RW, SA,

(54) Title: A MEDICAL THERAPEUTIC DEVICE




FIGURE 1

(57) Abstract: The present disclosure relates to the field of medicine and discloses a medical therapeutic device (100) for monitoring and treating medical condition of patients. The device (100) comprises an input unit (102), a plurality of sensors (104a-n), a control unit (106), a waveform generator unit (112), and a coupling means. The input unit (102) receives at least one input from a user. The sensors (104a-n) monitor a plurality of pre-determined parameters associated with the health of a patient generate detection signals based on the monitored parameters. The control unit (106) selects a program based on the input and the detection signals. The waveform generator unit (112) generates a therapeutic signal of pre-determined values of at least one of voltage, current, and frequency based on the selected program for facilitating treatment of medical condition corresponding to the selected program.



www.emedica.in

info.emedica@gmail.com

SC, SD, SE, SG, SK, SL, ST, SV, SY, TH, TJ, TM, TN, TR, TT, TZ, UA, UG, US, UZ, VC, VN, WS, ZA, ZM, ZW.

(84) Designated States (unless otherwise indicated, for every kind of regional protection available): ARIPO (BW, GH, GM, KE, LR, LS, MW, MZ, NA, RW, SD, SL, ST, SZ, TZ, UG, ZM, ZW), Eurasian (AM, AZ, BY, KG, KZ, RU, TJ, TM), European (AL, AT, BE, BG, CH, CY, CZ, DE, DK, EE, ES, FI, FR, GB, GR, HR, HU, IE, IS, IT, LT, LU, LV, MC, MK, MT, NL, NO, PL, PT, RO, RS, SE, SI, SK, SM, TR), OAPI (BF, BJ, CF, CG, CI, CM, GA, GN, GQ, GW, KM, ML, MR, NE, SN, TD, TG).

Declarations under Rule 4.17:


- as to the identity of the inventor (Rule 4.17(i))
- of inventorship (Rule 4.17(iv))

Published:

- with international search report (Art. 21(3))
- in black and white; the international application as filed contained color or greyscale and is available for download from PATENTSCOPE




PATH-LAB REPORTS OF THE PATIENTS





BEFORE & AFTER

BLOOD SUGAR LEVEL TRIALS AT SERUM INSTITUTE, PUNE



Report from Serum Institute


Before and After the Use of MediMagic

DIABETIS STUDY

Sr.	Name	Age	Date	Date	Age Date	BSL (G	lucome	ter) / mgdl	BSL -Random		Serum Insulin		
No				Pre	Post	After 2hrs	1	11	1	11	1		
1	Mr.Prakash D Kothawale	56	17/08/2018	123	94	92	126	96	5.9	3	1		
2	Mr.Sameer Joshi	45	17/08/2018	207	195	153	242	182	12.7	9.3	n		
3	Mr. Jaysing G Misal	61	17/08/2018	194	164	118	234	116	17.2	7.9	n		
4	Mr.Sunil Gaikwad	40	17/08/2018	240	179	134	264	162	16.4	6.6	1		
5	Mr.Aniket K.Tapkir	30	17/08/2018	214	210	188	238	220	5.4	6.1	1		
6	Mr.Milind Dhatrak	46	17/08/2018	243	177	106	88	102		41	1		
7	Mr.Abhijit Pawar	34	17/08/2018	102	100	83	243	86.3	12.2	7.8	1		

medication taken medication taken





Random BSL - BEFORE







Patient Name Age/Gender UHID/MR No

Visit ID

Mr. AMARJEET SINGH : 47 Y O M O D /M

: DWKA.0000000014 : DWKAOPV22 : Dr.SELF

Ref Doctor IP/OP NO

Collected Received Reported Status

: 07/Sep/2020 01:59PM : 07/Sep/2020 06:02PM : 07/Sep/2020 06:48PM

: Final Report : SL WAKAD

Client Name Patient location : WAKAD, Pune

2.7.				
DE	PARTMENT O	F BIOCHEMIST	RY	
Test Name	Result	Unit	Bio. Ref. Range	Method
GLUCOSE, RANDOM , SODIUM FLUORIDE PLASMA	449	mg/dL	70 - 140	Glucose oxidase
HBA1C, GLYCATED HEMOGLOBIN , WHOLE BLOOD-EDTA	11.4	%		HPLC
ESTIMATED AVERAGE GLUCOSE (eAG)	280	mg/dl		Calculated

Comment:

WHOLE BLOOD-EDTA

Reference Range as per American Diabetes Association (ADA):

REFERENCE GROUP	HBA1C IN %
NON DIABETIC ADULTS > 18 YEARS	<5.7
AT RISK (PREDIABETES)	5.7 – 6.4
DIAGNOSING DIABETES	>= 6.5
DIABETICS	No. 10 Personal Section 1
EXCELLENT CONTROL	6-7
FAIR TO GOOD CONTROL	7 – 8
UNSATISFACTORY CONTROL	8 – 10
POOR CONTROL	>10

Note: Dietary preparation or fasting is not required.

- 1. AIC test should be performed at least two times a year in patients who are meeting treatment goals (and who have stable glycemic control).
- 2. Lowering A1C to below or around 7% has been shown to reduce microvascular and neuropathic complications of type 1 and type 2 diabetes. When mean annual HbA1c is <1.1 times ULN (upper limit of normal), renal and retinal complications are rare, but complications occur in >70% of cases when HbA1c is >1.7 times ULN.
- 3. Falsely low HbA1c (below 4%) may be observed in patients with clinical conditions that shorten erythrocyte life span or decrease mean erythrocyte age. HbA1c may not accurately reflect glycemic control when clinical conditions that affect erythrocyte survival are present. Fructosamine may be used as an alternate measurement of glycemic control



www.emedica.in ■ info.emedica@gmail.com

Random BSL - AFTER







Patient Name : Mr.AMARJEET SINGH Age/Gender : 47 Y O M O D /M UHID/MR No : DWKA.0000000014 Visit ID : DWKAOPV26

: Dr.SELF Ref Doctor

Collected : 07/Sep/2020 05:16PM Received : 07/Sep/2020 08:36PM : 07/Sep/2020 08:52PM Reported Status : Final Report

: SL WAKAD Client Name Patient location : WAKAD, Pune

IP/OP NO :		Patient location	: WAKAD,Pune	
t	DEPARTMENT O	F BIOCHEMIST	RY	
Test Name	Result	Unit	Bio. Ref. Range	Method
GLUCOSE, RANDOM, SODIUM FLUORIDE	308	mg/dL	70-140	HEXOKINASE



Page 3 of 4



Random BSL - BEFORE







: Mrs.JYOTI KALSI Patient Name Age/Gender : 47 Y O M O D /F UHID/MR No : DWKA.0000000015 : DWKAOPV23 Visit ID Ref Doctor : Dr.SELF

Collected : 07/Sep/2020 02:03PM Received : 07/Sep/2020 06:02PM Reported : 08/Sep/2020 02:58PM Status : Final Report

Client Name : SL WAKAD : WAKAD, Pune Patient location

DEPARTMENT OF BIOCHEMISTRY					
Test Name	Result	Unit	Bio. Ref. Range	Method	
GLUCOSE, RANDOM, SODIUM FLUORIDE PLASMA	329	mg/dL	70-140	HEXOKINASE	
UDALE CUMELTER UTALOGUARIN	0.5			LIBI C	

HBA1C, GLYCATED HEMOGLOBIN, WHOLE BLOOD-EDTA	9.6	%	HPLC
ESTIMATED AVERAGE GLUCOSE (eAG), WHOLE BLOOD-EDTA	229	mg/dL	Calculated

IP/OP NO

Reference Range as per American Diabetes Association (ADA):

REFERENCE GROUP	HBA1C IN %
NON DIABETIC ADULTS > 18 YEARS	<5.7
AT RISK (PREDIABETES)	5.7 – 6.4
DIAGNOSING DIABETES	>= 6.5
DIABETICS	N. LEVICE MICHIGAN
- EXCELLENT CONTROL	6-7
FAIR TO GOOD CONTROL	7 – 8
UNSATISFACTORY CONTROL	8 – 10
POOR CONTROL	>10


Note: Dietary preparation or fasting is not required.

- 1. AIC test should be performed at least two times a year in patients who are meeting treatment goals (and who have stable glycemic control).
- 2. Lowering A1C to below or around 7% has been shown to reduce microvascular and neuropathic complications of type 1 and type 2 diabetes. When mean annual HbA1c is <1.1 times ULN (upper limit of normal), renal and retinal complications are rare, but complications occur in >70% of cases when HbA1c is >1.7 times ULN.
- 3. Falsely low HbA1c (below 4%) may be observed in patients with clinical conditions that shorten erythrocyte life span or decrease mean erythrocyte age. HbA1c may not accurately reflect glycemic control when clinical conditions that affect erythrocyte survival are present. Fructosamine may be used as an alternate measurement of glycemic control



Apollo Health and Lifestyle Limited





Random BSL - AFTER







: Mrs.JYOTI KALSI : 47 Y 0 M 0 D /F Patient Name Age/Gender UHID/MR No : DWKA.0000000016 Visit ID : DWKAOPV27

Ref Doctor : Dr.SELF

IP/OP NO

Collected : 07/Sep/2020 05:27PM Received : 07/Sep/2020 08:36PM Reported : 07/Sep/2020 08:52PM

: Final Report Status Client Name : SL WAKAD Patient location : WAKAD, Pune

DEPARTMENT OF BIOCHEMISTRY						
Test Name	Result	Unit	Bio. Ref. Range	Method		
GLUCOSE, RANDOM, SODIUM FLUORIDE PLASMA	207	mg/dL	70-140	HEXOKINASE		



45



@ www.emedica.in

Random BSL - BEFORE



ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

• Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur. • Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Vaidehi Tannirwar

Age/Sex :- Female

Ref. By :-

Self

Date :- 20/7/19

SAMPLE NUMBER :-346

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random (Pre- Trail)

152 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

INVESTIGATION

NORML RANGE

Insulin (Pre- Trail)

26.7 uU/mL - 0 - 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



Random BSL - AFTER



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Vaidehi Tannirwar

Age/Sex :- Female

Ref. By :-

Self

349 (Post - Trail)

Date :- 20/7/19

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random (Post-Trail)

SAMPLE NUMBER :-

107 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

Insulin (Post - Trail)

RESULT

NORML RANGE

14.7 uU/mL

0 -- 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



Random BSL - BEFORE



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Nandini Paithankar

· Age/Sex :- Female

Ref. By :- Se

SAMPLE NUMBER :- 371 (Pre - Trial)

Date :- 21/7/19

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random

227 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

INVESTIGATION

RESULT

NORML RANGE

Insulin

47.8 uU/mL

0 -- 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.

Timing: MIDC Morning 9.00 A.M. to 2.30 P.M. Eve.: 6.30 P.M. to 8,15 P.M. Sunday: 9.00 A.M. to 12.00 P.M.



@ www.emedica.in

Random BSL - AFTER



Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

 Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur. • Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Nandini Paithankar

Age/Sex :- Female

Ref. By :- . Self

SAMPLE NUMBER :- 374 (Post - Trial)

Date :- 21/7/19

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random

173 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

INVESTIGATION

RESULT

NORML RANGE

24.4 uU/mL

0 -- 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.

Timing: MIDC Morning 9.00 A.M. to 2.30 P.M. Eve.: 6.30 P.M. to 8.15 P.M. Sunday: 9.00 A.M. to 12.00 P.M:



Random BSL - BEFORE



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C'o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.

 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Rachana Jain

Age/Sex :- Female

Ref. By :-

SAMPLE NUMBER :-

Self

370 (Pre - Trial)

Date :- 21/7/19

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random

167 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

INVESTIGATION

RESULT

NORML RANGE

Insulin

20.7 uU/mL

0 - 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



Random BSL - AFTER



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Rachana Jain

. Age/Sex :- Female

Ref. By :-

SAMPLE NUMBER :-

Self

373 (Post - Trial)

Date :- 21/7/19

REPORT ON BIOCHEMISTRY

INVESTIGATION

RESULTS

NORMAL RANGE

Blood Glucose Random

146 mg/dl.

Less than 160 mg/dl.

REPORT ON HARMONE ESTIMATION

INVESTIGATION -

RESULT

NORML RANGE

Insulin

18.6 uU/mL

0 -- 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



Random Cholesterol **BEFORE**



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Dr. Inder Gundecha

Age/Sex :- Male

Ref. By :-

Self

Date :- 20/7/19

SAMPLE NUMBER :- 357 (Pre-Trail)

REPORT ON BIOCHEMISTRY

INVESTIGATION	RESULTS	NORMAL RANGE
Blood Glucose Random	87 mg/dl.	Less than 160 mg/dl.
LIPID PROFILE		
Cholesterol Total	360 mg/dl.	130-250 mg/dl.
HDL	48 mg/dl.	30-70 mg/dl.
L.DL	288 mg/dl.	Upto 155 mg/dl.
VLDL	27 mg/dl.	Upto 35 mg/dl.
Şr. Triglycerides	135 mg/dl.	Upto 170 mg/dl.
Chol./HDL Ratio	7.5	Upto 4.5

^{*} Done on ion selective Analyzer.

REPORT ON HARMONE ESTIMATION

INVESTIGATION

RESULT

NORML RANGE

Insulin

16.1 uU/mL

0 -- 24 uU/mL ·

Dr. Pravis Cadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.

Timing: MIDC Morning 9.00 A.M. to 2.30 P.M. Eve.: 6.30 P.M. to 8.15 P.M.



Random Cholesterol **AFTER**



GADKARI PATHOLOGY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :- Dr. Inder Gundecha

Age/Sex :- Male

Ref. By :-

SAMPLE NUMBER :-

Self

361 (Post - Trail)

Date :- 20/7/19

dl.

REPORT ON BIOCHEMISTRY

LIVESTIGATION	RESULIS	NORMAL RANGE
Blood Glucose Random	74 mg/dl.	Less than 160 mg/e
LIPID PROFILE		
Cholesterol Total	195 mg/dl.	130-250 mg/dl.
HDL.	52 mg/dl.	30-70 mg/dl.
LDL	125 mg/dl.	Upto 155 mg/dl.
VLDL	18 mg/dl.	Upto 35 mg/dl.
Sr. Triglycerides	92 mg/dl.	Upto 170 mg/dl.
Chol./HDL Ratio	3.75	Upto 4.5

Done on ion elective Analyzer.

REPORT ON HARMONE ESTIMATION

 Investigation
 RESULT
 NORML RANGE

 Insulin
 7.5 uU/mL
 0 - 24 uU/mL

Dr. .Pravin Gadkari MD -(Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.

Timing: MIDC Morning 9.00 A.M. to 2.30 P.M. Eve.: 6.30 P.M. to 8.15 P.M.



Random Cholesterol **BEFORE**



GADKARI PATHOLOGY ELISA & MICRO BIOLOGY LABORATORY

Dr. Pravit Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name :-

Mrs. Amisha Gundecha

Age/Sex :- Female

Ref. By :-

TREETERS A PLOSE

Self

Date :- 20/7/19

NORMAL PANCE

SAMPLE NUMBER :- 358 (Pre-Trail)

REPORT ON BIOCHEMISTRY

DECIII TO

INVESTIGATION	RESULIS	NORMAL RANGE
Blood Glucose Random LIPID PROFILE	101 mg/dl.	Less than 160 mg/dl.
Cholesterol Total	281 mg/dl.	130-250 mg/dl.
HDL	40 mg/dl.	30-70 mg/dl.
LDL	196 mg/dl.	Upto 155 mg/dl.
VLDL	45 mg/dl.	Upto 35 mg/dl.
Sr. Triglycerides	223 mg/dl.	Upto 170 mg/dl.
Chol./HDL Ratio	7.0	Upto 4.5

^{*} Done on ion selective Analyzer. .

REPORT ON HARMONE ESTIMATION

INVESTIGATION	RESULT	NORML RANGE		
Insulin	43.3 uU/mL	0 24 uU/mL		

Dr. Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



Random Cholesterol **AFTER**



GADKARI PATHOLOGY

Dr. Pravin Gadkari M.B.B.S., MD (Path) Director NRPL & Ayush Blood Bank

Shop No. 26, Upper Ground Floor, Madhu Madhav Towers, Laxmi Bhavan Square, Dharampeth, Nagpur - 10
 Branch: C/o. Dr. Dhondse Hospital, Radke Layout, Balaji Nagar, Hingna Road, MIDC, Nagpur.
 Tel. No.: 8237067621, 8087045795 | E-mail: gadkaripravin@yahoo.co.in

Name:-

Mrs. Amisha Gundecha

Age/Sex :- Female

Ref. By :-

Self

IT

Date :- 20/7/19

SAMPLE NUMBER :- 360 (Post - Trail)

REPORT ON BIOCHEMISTRY

INVESTIGATION	RESULTS	NORMAL RANGE
Blood Glucose Random	91 mg/dl.	Less than 160 mg/dl
LIPID PROFILE		
Cholesterol Total	234 mg/dl.	130-250 mg/dl.
HDL	48 mg/dl.	30-70 mg/dl.
LDL	156 mg/dl.	Upto 155 mg/dl.
VLDL	30 mg/dl.	Upto 35 mg/dl.
Sr. Triglycerides	151 mg/dl.	Upto 170 mg/dl.
Chol./HDL Ratio	4.87	Upto 4.5

^{*} Done on ion selective Analyzer.

REPORT ON HARMONE ESTIMATION

Insulin 22.8 uU/mL 0 -- 24 uU/mL

Dr. .Pravin Gadkari MD (Path)

Timing: 8.00 A.M. to 9.00 P.M.

Sunday: 8.00 A.M. to 1.00 P.M.



@ www.emedica.in ■ info.emedica@gmail.com

Random Cholesterol BEFORE



IP/OP NO





Patient Name Age/Gender : 38 Y 0 M 0 D /M UHID/MR No : DPVP.0000000718 Visit ID : DPVPOPV1096 Ref Doctor : Dr.SELF

Collected : 24/Aug/2018 02:26PM Received : 24/Aug/2018 05:14PM Reported : 24/Aug/2018 05:33PM : Final Report Status Client Name : PCC VADGAONSHERI PUNE

Client Code : PCC0157 **DEPARTMENT OF BIOCHEMISTRY**

Test Name	Result	Unit	Bio. Re	f. Range	Method
LIPID PROFILE, SERUM					
TOTAL CHOLESTEROL	269	mg/dL	<200	Enzymatio	(CHE/CHO/POD)
TRIGLYCERIDES	258	mg/dL	<150	Enzymatic	(Lipase/GK/GPO/POD
HDL CHOLESTEROL	39	mg/dL	40-60	THE REAL PROPERTY.	asure PEG
NON-HOL CHOLESTEROL	230	mg/dL	<130	Calculates	d
LDL CHOLESTEROL	178.4	mg/dL	<100	Calculates	5
VLDL CHOLESTEROL	51.6	mg/dL	<30	Calculated	4
CHOL/HOL RATIO	6.90		0.4.97	Calculates	4

Reference Interval as per National Cholesterol Education Program (NCEP) Adult Treatment Panel III Report.

	Desirable	Borderline High	High	Very High
TOTAL CHOLESTEROL	< 200	200 - 239	≥ 240	
TRIGLYCERIDES	<150	150 - 199	200 - 499	≥500
LDL	Optimal < 100 Near Optimal 100-129	130 - 159	160 - 189	≥ 190
HDL	≥60			
NON-HDL CHOLESTEROL	Optimal <130; Above Optimal 130-159	160-189	190-219	>220

Measurements in the same patient can show physiological and analytical variations.

NCEP ATP III identifies non-HDL cholesterol as a secondary target of therapy in persons with high triglycendes.

*** End Of Report ***





Page 1 of 1

SIN No.BIO 1018788
This test has been performed at Apolto Health and Lifestyle Ltd Reference Regional Lab, Pone

Apollo Health and Lifestyle Limited

(CNI- LIES I USBADOOM, COHOODIS Riegd, Office: 1978/shop Gardient, R.A. Pursan, Chemiul 400 028, Tamil Riadis, India: Email ICt. Infortagolishiczen

www.apullodlagnestics.in



Random Cholesterol **AFTER**

PATIENT NAME : Mr. Jayendra Singh

AGE: 38 Yrs / M

REFFERED BY: Dr. Self

NHDL CHOLESTEROL

DATE: 25-08-2018

BIOCHEMISTRY LIPID PROFILE

Test Name	Value	Unit	Reference Range
T.CHOLESTEROL	206	mg/dl	Upto 200
S.TRIGLYCERIDES	146	mg/dl	25-200
HDL CHOLESTEROL	41	mg/dl	M:30-70; F:35-90
VLDI.	29.2	mg/dl	5-40
LDL CHOLESTEROL	135.8	mg/dl	85-130
T.CHOLESTEROL/HDL	5.02	Ratio	3.0-5.0
LDL/HDL	3.31	Ratio	1.5-3.5

mg/dl

This is only a professional opinion & not a diagnosis. Please Correlate with clinical conditions. Alert !!! 10-12 hours fasting is mandatory for lipid parameters. If not, values might fluctuate,

DR.SAMIR AGARWAL DNB (PATH)

CONSULTANT PATHOLOGIST



HbA1C- BEFORE







Patient Name : FRANCIS DOMINIC FERNANDEZ Receiving Branch : SC Taman Megah

Identification : S2751828A Age / Gender : 65 year - Male Barcode : D822359912 (EC6038145)

Date of Birth : 06/11/1957 Nationality : Singaporean Refer/Policy No : Client Name : Walk-in - WALK-IN - Walk-in Specimen :

 Doctor's Name
 : Dr Tan Hai Ning
 Collected
 : 17/11/2022 09:35 AM

 Panel Clinic
 : Dr Tan Hai Ning Received
 : 17/11/2022 06:20 PM

 Report ready
 : 18/11/2022 10:34 AM

TEST NAME	RESULT	FLAG	UNIT	REFERENCE	METHOD
HBA1C (糖化血紅蛋白)					
HbA1c (NGSP)	8.3	н	%	Normal: < 5.7 Pre-diabetes: 5.7 - 6.2 Diabetes: ≥ 6.3	Enzymatic
HbA1c (IFCC)	67	н	mmol/mol	Normal: < 39 Pre-diabetes: 39 - 44 Diabetes: ≥ 45 (Management of T2DM CPG, 6th Ed.)	Calculated

Validated By: NG SHI MAN - BP CLINICAL LAB SDN BHD (GLENMARIE)
Final Report Verified By: SHAHIRAH AZZAHARAH ZURAIMI - BP CLINICAL LAB SDN BHD (TAMAN MEGAH)

Page 1 of 1

Disclaimer: The above reference ranges are based on adult population unless otherwise stated. [^] indicates age-specific paediatric reference range.

For diagnostic purposes, all laboratory test results should be used in conjunction with clinical evidence.

For further confirmation, please repeat test with another fresh speciment, if desired.

Should you have further enquiries, please contact your nearest BP branch or mail us at bphq@bphealthcare.com

BP Healthcare Group @ Glenmarie, Temasya @ Glenmarie, Jalan Pendaftar U1/54, Section U1, Shah Alam, Selangor, Malaysia



@ www.emedica.in

HbA1C- AFTER



PATHOLOGY REPORT

B. P. CLINICAL LAB SDN. BHD. (152314-H)





Patient Name

Client Name

: FRANCIS DOMINIC FERNANDEZ

Identification : S2751828A Date of Birth : 06/11/1957 Age / Gender : 65 year - Male

Nationality : Singaporean

: Walk-in - WALK-IN - Walk-in

Doctor's Name : Dr Koay Weng Kuan Panel Clinic : Dr Koay Weng Kuan

: SC Taman Megah

: D822412740 (EC6099282) Barcode

Refer/Policy No

Receiving Branch

Specimen : Fasting

Collected : 17/12/2022 10:34 AM Received : 17/12/2022 02:21 PM Report ready : 18/12/2022 12:13 PM

TEST NAME	RESULT	FLAG	UNIT	REFERENCE	METHOD
HBA1C (糖化血紅蛋白)					
HbA1c (NGSP)	7.7	Н	%	Normal: < 5.7 Pre-diabetes: 5.7 - 6.2 Diabetes: ≥ 6.3	Enzymatic
HbA1c (IFCC)	61	н	mmol/mol	Normal: < 39 Pre-diabetes: 39 - 44 Diabetes: ≥ 45 (Management of T2DM CPG, 6th Ed.)	Calculated
Glucose (Plasma) (血糖)	6.8		mmol/L	Fasting: < 6.1 Random: < 7.8	Hexokinase

Validated By: NURUL NADHIRAH BINTI JOHARI - BP CLINICAL LAB SDN BHD (GLENMARIE) Final Report Verified By: SHAHIRAH AZZAHARAH ZURAIMI - BP CLINICAL LAB SDN BHD (TAMAN MEGAH)

Page 1 of 1

Disclaimer: The above reference ranges are based on adult population unless otherwise stated. [^] indicates age-specific paediatric reference range.

For diagnostic purposes, all laboratory test results should be used in conjunction with clinical evidence. For further confirmation, please repeat test with a nother fresh speciment, if desired. Should you have further enquiries, please contact your nearest BP branch or mail us at bphq@bphealthcare.com BP Healthcare Group @ Glenmarie, Temasya @ Glenmarie, Jalan Pendaftar U1/54, Section U1, Shah Alam, Selangor, Malaysia



Test Name

@ www.emedica.in

HbA1C- BEFORE

Name Mr. Deepak Moghe Ref. By Dr. Mrs. Mardikar MD Sample Rec/Coll. Dt 25/05/2020 Sample Rec/Coll Time 7:17:00 AM

58 Yrs. M / Age / Sex Sample ID No.

Type of Sample

EDTA Blood, EDTA BLOOD FASTING, Seru

Recd. From Outside Yes

HAEMATOLOGY

Test Name Value Unit Biological Ref. Interval

HbAIC (GLYCOSYLATED HAEMOGLOBIN)

Investigation Interpretation Results HbA1c % :-11.45 % Non diabetic : 4.3 - 5.7 %

Prediabetic: 5.7-6.3 % Good diabetic control : 6.3 - 7.3 %

Fair control : 7.3 - 8.3 % Poor control : 8.3 % & above

Method :- Nephelometry & Turbidometry Done on MISPA -i2

Note: - * HbA1c (Or GlycoHb) is made by post synthetic modification of Haemoglobin A at a slow rate directly dependant on blood glucose concentration during the 120 day life span of RBC.HbA1c levels may double or even triple in diabetics, depending on level of hyperglycemia, and correlate well with control of diabetes.

BIOCHEMISTRY


Biochemistry done on VITROS 250 Dry Chemistry Analyzer

Test Name	Value	Unit	Biological Ref. Interval
BLOOD SUGAR FASTING	173	mg/dl	74 - 106
BLOOD SUGAR FASTING	160	mg/dl	74 - 106
	IMMUNO	LOGY	

Unit

Contd...2

Biological Ref. Interval





HbA1C- AFTER

Sample Rec/Coll. Dt 14/06/2020 Sample Rec/Coll Time 12:50:07

LIVER FUNCTION TEST (LFT)

Type of Sample EDTA Blood, EDTA BLOOD RANDAM, F

Recd. From Outside Yes

Test Name	Value	Unit	Biological Ref. Interval
VITROS 250 Dry Chem. Analyzer			
TOTAL PROTEIN	8.6	gm/dl	6.3 - 8.2
ALBUMIN	4.70	gm/dl	3.5 - 5.0
GLOBULIN*	3.90	gm/dl	2.0 - 4.0
SGOT	26	U/L	17 - 59
SGPT	34	U/L	21 - 72
ALKALINE PHOSPHATASE	83	U/L	38 - 126
BILIRUBIN TOTAL	0.7	mg/dl	0.2 - 1.3
DIRECT BILIRUBIN (BC)	0.4	mg/dl	0.0 - 0.30
INDIRECT BILIRUBIN (BU)	0.30	mg/dl	0.1 - 1.1

LFT INTERPRET

HAEMATOLOGY

Hbaic (GLYCOSYLATED HAEMOGLOBIN)

Investigation Results Interpretation

HbA1c %:- 9.0 %

Non diabetic: 4.3 - 5.7 %
Prediabetic: 5.7-6.3 %
Good diabetic control: 6.3 - 7.3

Fair control : 7.3 - 8.3 % Poor control : 8.3 % & above

Method: HPLC (Done By HB-VARIO from ERBA)

Contd...4



www.emedica.in ■ info.emedica@gmail.com

HbA1C- BEFORE



64





LABORATORY REPORT

: PASTOR. LEO QUADROS eferred By : GENERAL(MEDICAL) ge ; 69 Year(s) 10 Month(s) 1 Day(s) rdered Date : 09-Feb-2022 07:44:56

Report Verified

ollected Date : 09-Feb-2022 07:50:50

MRD No. : AA733326

: OPORD0002628769 Lab No

: Male Sex

Reported on : 09-Feb-2022 12:10:03

Test	Result	Units	Biological Reference Interva
LDL:HDL Ratio	2	Ratio	0 - 3.5
Report Verified Blood Glucose PPBS	Yes		
Specimen : Plasma2hr			
2 hour Post Prandial (PP)Blood glucose Method : Hexokinase	331	mg/dL	80 - 140
Report Verified	Yes		
BA1C			
Glycosylated Haemoglobin (HBA1c)			
Specimen : EDTA Blood Sample			
Glycosylated Haemoglobin (HbA1c)	9.1	%	Normal - Less than 5.7% Pre-diabetes - 5.7 - 6.4% Diabetes : 6.5% and higher

Yes

zassel-DECREASED AND STREET BY STREET

For Home Collection Call: 080 22024450 / 9480689775



HbA1C- AFTER





 Visit No.
 : 2223AB00052873

 Name
 : Mr. LEO QUADROS

 Age / Sex
 : 70Years / Male

 Ref Dr
 : Dr. SELF

 Reported On
 : 15-Dec-2022

 1:16 AM

HbA1C

	HOATC		Washington and American
INVESTIGATIONS	RESULT	REF. RANGE	UNITS
GLYCOSYLATED HEMOGLOBIN (HBA1c)	7.90	Ref range Poor control : > 7.6% Fair Control : 6.8 - 7.6 % Good Dlabetic Control : 5	.5 - 6.8%
Method : HPLC		Normal : 4 .2 - 6.2%	
MEAN BLOOD GLUCOSE	180	Mean Average Blood Gluc Excellent Control : 90 - 12 dL Good Control: 121 - 150	20 mg /
		Average Control : 151 - 18 dL Action Suggested : 181 - 2 dL	210 mg/
		Panic Value : >211 mg / dl	L
Biorad D-10			

Checked/- SUDHAKAR

Mrs. Mani G., MSC Microbiologist Natarajan.m

Dr. M. Natarajan, MBBS, MD

Consultant Pathologist KMC NO : 34875

 $Diagnostic\ Centre: \#83,\ N.K.\ Complex,\ Kammanahalli\ Main\ Road,\ Bangalore-84,\ Tel: 080\ 25478934/65983501/25805119$

Report relates to the sample tested. All reported results are for the information of the referring doctor only. Any difficulty in interpreting the results, kindly contact the signatory.



@ www.emedica.in

HbA1C- BEFORE

PROCESSED AT :

Thyrocare, Sohrabh Hall, 112, A Wing, 1st Floor, Sangamwadi, Pune - 411 001 I hyrocare

The Trust. The Truth.

Corporate Office: Thyrocare Technologies Limited ♥ D-37/3, TTC MIDC, Turbhe, Navi Mumbai - 400703

NAME

: RAJKUMAR HEMDEV(79Y/M)

REF. BY

TEST NAME

: AAROGYAM B **TEST ASKED**

SAMPLE COLLECTED AT :

FA 2 YUTOPIA COMPLEX OPPOSITE RAHEJA

GARDEN WANWORI PUNE ,

TECHNOLOGY

VALUE 10.2

UNITS

mg/dl

HbA1c - (HPLC - NGSP Certified)

H.P.L.C

Reference Range:

Reference Range: As per ADA Guidelines

Below 5.7% : Normal 5.7% - 6.4% : Prediabetic

: Diabetic >=6.5%

Guidance For Known Diabetics

Below 6.5% : Good Control 6.5% - 7% : Fair Control

7.0% - 8% : Unsatisfactory Control

246

: Poor Control >8%

Method: Fully Automated H.P.L.C. using Biorad Variant II Turbo, NGSP Certified. CALCULATED

AVERAGE BLOOD GLUCOSE (ABG)

Reference Range:

90 - 120 mg/dl : Good Control

121 - 150 mg/dl : Fair Control

151 - 180 mg/dl : Unsatisfactory Control

: Poor Control > 180 mg/dl

Method: Derived from HBA1c values

Please correlate with clinical conditions.

Sample Collected on (SCT)

Sample Received on (SRT)

Report Released on (RRT)

Sample Type

Barcode

Labcode

:0610043934/AE847

.06 Oct 2020 07:56

.06 Oct 2020 17:23

. 06 Oct 2020 18:43

:R1402709

Dr. Prachi Sinkar MD(Path)

Page: 7 of 9



HbA1C- AFTER



Checked By Technician

K.E.M. HOSPITAL, PUNE - 11.

489, Rasta Peth, Sardar Moodliar Road. TEL.:020-66037300, E-mail : diagnosticsmanager@kemhospital.org



Certificate No.: M0362 NABL Accredited Laboratory

: Mr. RAJKUMAR LALCHAND HEMDEV PRN : 780262 IP No.

Patient : Mr. RAJKUMAR LALCHAND HEMDEV PKN

Age/Sex : 79 Yrs/Male Visit No. : OP-7 Date : 01/06/2021

Referred By : Dr. Smita Dhadge Sample Collected : 01/06/2021 12:43

Location : OPD Sample Recvd in Lab : 01/06/2021 12:43

Location : OPD Sample Recvd in Lab Sample Recv

Specimen : Blood : Final

Lab No. : 0032115221 Status : Final

DEPARTMENT OF ENDOCRINOLOGY

Bio, Ref. Interval

Investigations Result Bio, Ref. Interval

GLYCOSYLATED HB (HB AIC) 7.900 % NON DIABETES LESS THAN 5.7 %

PRE DIABETES: 5.7 - 6.4 %
DIABETES: 6.5 % OR HIGHER

Method: HPLC _____End Of Report-----

DR. SADANAND NAIK
M.Sc.Ph.D.(Med.Biochemistry)
HOD-Clinical Biochemistry


Entered By:thp9686



HbA1C- BEFORE









HbA1C- AFTER





Creatinine - BEFORE

Medical Laboratory Report

35

LUPIN DIAGN STICS

 Patient Information

 Name
 :
 Mr.ASHOK SAINI

 Age/Gender
 :
 72Y 2M 3D/Male

 MobileNo
 :
 9892697751

 UHID
 :
 LDAA00031381

 Address
 :
 1102 tulip, dosti acre,

: 1102 tulip, dosti acre, behind antop hill bus depot. 400037 Report Status : Final Report

Client/Doctor Information Ref Doctor : Dr.SELF Client Code : LMPCC0121 Client Add. : WADALA EAS Client No. :

Client Name : PCC WADALA DOSTI

ACRES

Test Name Result Bio. Ref. Range Unit Method

Blood Urea Nitrogen (BUN), SERUM

Blood Urea Nitrogen (BUN)	57.4	8.0-23.0	mg/dl	Kinetic Urease
Result Rechecked				

Creatinine, Serum

Creatinine	6.78	0.70-1.20	mg/dL	Kinetic Jaffe
Result Rechecked				

Electrolytes, Serum

Sodium	140.2	136 - 145	mmol/L	Ion-selective electrode
Potassium	5.7	3.4 - 5.1	mmol/L	Ion-selective electrode
Chloride	105.5	98-107	mmol/L	Ion-selective Electrode

Page 1 of 2





This test has been performed at Lupin Diagnostics Laboratory, NRL MUMBAI National Reference Laboratory, Plot No.C-533, MIDC, TTC Industrial Area, Pawane



www.emedica.in ■ info.emedica@gmail.com

Creatinine - AFTER

: LDAL715

: 09/Jun/2022 09:42

Patient Information Name : Mr.ASHOK SAINI Age/Gender: 71Y 3M 9D/Male MobileNo : 9892697751 UHID : LDAA00039220

Address : 1102 Dosti Tulip Dosti Acres

Reported : 09/Jun/2022 13:49 IP/OP/Barcode:

Visit ID

Collected

Received

Specimen Information

Report Status : Final Report

: 09/Jun/2022 12:17 Client Add. : WADALA EAS Client No.

Client Name : PCC WADALA DOSTI ACRES

Client/ Doctor Information

Client Code : LMPCC0121

Ref Doctor : Dr.SELF

Test Name	Result	Bio. Ref. Range	Unit	Method
Blood Urea Nitrogen (BUN) , SERUM				

Creatinine, Serum

6.07	0.70-1.20	mg/dL	Kinetic Jaffe
	6.07	6.07 0.70-1.20	6.07 0.70-1.20 mg/dL


Electrolytes, Serum

Sodium	134.0	136 - 145	mmol/L	Ion-selective electrode
Potassium	5.2	3.4 - 5.1	mmol/L	Ion-selective electrode
Chloride	106.9	98-107	mmol/L	Ion-selective Electrode

Page 1 of 2











Creatinine - BEFORE





Creatinine - AFTER






@ www.emedica.in

Creatinine - BEFORE

Medicare Hospital & Research Centre

4/5, Ravindra Nagar, Old Palasia, INDORE-452 018 (M.P.) Ph.:0731-4271600, 2490577, 2492621, 4065616-17-18 • Fax: 4266243 Mobile: 88899-12611, E-mail: medicarehospitalindore@gmail.com Website: www.medicarehospitalindore.com





Patient Name :

Mrs. UMA SARAF [MRN-210800808]

Age / Gender :

72 Yr / Female

Address :

BEEMA NAGAR, Indore, MADHYA PRADESH

Req. Doctor: Self

Regn. ID:

WALKIN.21-22-5396

BIOCHEMISTRY

RESULT

Request Date : Sample No. :

INVESTIGATIONS

09-08-2021 11:08 AM

09-08-2021 11:56 AM Acceptance Date :

BI2442

Reporting Date: 09-08-2021 12:40 PM

Reporting Status: Finalized

CREATININE

4.9 mg/dL *

REFERENCE RANGE 0.50 - 1.40 mg/dL

BLOOD UREA NITROGEN *[SERUM]

Blood Urea Blood Urea Nitrogen (BUN) POTASSIUM.

92 mg/dL * 42.99 mg/dL + 20.00 - 40.00 mg/dL 6.00 - 20.00 mg/dL

5.2 Meg/L *

3.50 - 5.00 Meg/L

END OF REPORT.

KSRaj

Dr. Karamchand Rajani D.C.P (London)

Dr. Chandrashekhar (MD, Pathology)

Dr. Pradeep Apte (M.Sc,DSM,Ph.D)



Page 1 of 1



@ www.emedica.in info.emedica@gmail.com
 info.emedica@gmail.com

Creatinine - AFTER

Medicare Hospital & Research Centre

4/5, Ravindra Nagar, Old Palasia, INDORE-452 018 (M.P.) Ph.:0731-4271600, 2490577, 2492621, 4065616-17-18 • Fax: 4266243

Mobile: 88899-12611, E-mail: medicarehospitalindore@gmail.com

Website: www.medicarehospitalindore.com





Patient Name :

Mrs. UMA SARAF [MRN-210800808]

Age / Gender :

72 Yr / Female

Address:

BEEMA NAGAR, Indore, MADHYA PRADESH

Req. Doctor:

SELF

Regn. ID:

WALKIN.21-22-17438

BIOCHEMISTRY

Request Date : Sample No. :

20-10-2021 06:51 AM

BI6961

INVESTIGATIONS

20-10-2021 11:20 AM

Reporting Date: 20-10-2021 11:20 AM

Reporting Status: Finalized

Acceptance Date :

REFERENCE RANGE

CREATININE

4.3 mg/dL *

0.50 - 1.40 mg/dL

GLUCOSE (FASTING)

76 mgm%

RESULT

70.00 - 110.00 mgm%

END OF REPORT.

KS Ray -


Dr. Karamchand Rajani D.C.P. (London)

Dr. Chandrashekhar (MD, Pathology)

Dr. Pradeep Apte (M.Sc,DSM,Ph.D)




Page 1 of 1






Creatinine - BEFORE




E-mail: visiondiagnosticss@gmail.com, Ph: 080-2340 4871, Mob: 98807 82516






Creatinine - AFTER





Thyroid - BEFORE





Thyroid - AFTER

38211050006

Bill No. : 8747

Patient Name

MRS. MONIKA S. DESARDA

Age/Sex : 40 / Female

Center : DIAGNOPEIN, BIBWEWADI

Referred By. : SELF

PID : 31137

Registration : 15/04/21, 09:01 AM Reported : 15/04/21, 01:44 PM

Status : Final Report

Investigation Result Units Bio.Ref. Interval

IMMUNOASSAY

TFT - THYROID FUNCTION TEST

 Tri-iodothyronine, Total (T3)
 1.173
 ng/ml
 0.60 - 2.15

 Thyroxine, Total (T4)
 93.59
 ng/ml
 52 - 127

 Method : (Serum, CLIA)
 Ultra TSH
 3.824
 ulU/mL
 0.3 - 4.5

Method : (Serum, CLIA)

Note

Reference ranges are age dependent and somewhat method dependent, the above values are approximate expected values. Reference range for Pregnant women are as per American Thyroid Association guidelines. T3 & T4 levels may be abnormal in euthyroid sick syndrome. Physiological rise in Total T3 / T4 levels is seen in pregnancy and in patients on steroid therapy. TSH may be the only parameter abnormal in subclinical hypothyroidism. Many drugs and intercurrent illnesses are associated with alterations in serum TSH. Hence all results should be clinically correlated and if required repeated when patient is clinically stable.

END OF REPORT

Checked by - Prasad

Dr. Payal Kalwar M.D. Pathologist



Heart - BEFORE



NOBLE HOSPITALS, PUNE

South East Pune's 1st NABH Accredited Hospital

153, Magarpatta City Road, Hadapsar, Pune 411 013. Tel.: 020-8628 5000 Fax : 020 - 6628 5199 CIN NO. U85110PN1996PTC103171, Visit Us at : www.noblehospitalspune.com



DEPARTMENT OF CARDIOLOGY

10th FEB, 2021

MRS.BHUTADA SHOBHA.

CORONARY ANGIOGRAPHY REPORT.

LEFT MAIN

: Distal 60 % stenosis.

LEFT ANTERIOR DESCENDING CORONARY : Mid LAD 85-90 % stenosis. diffuse disease poor target.

LEFT CIRCUMFLEX CORONARY ARTERY.

: OM,90 % calcific stenosis. Moderately calcified.

RIGHT CORONARY

: Proximal 60 % stenosis. Mid RCA diffuse disease RPDA 90 % stenosis.

RECOMMENDATION

: Surgical opinion.

Col Dr. RAHUL SAWANT MD MRCP (UK) CCT (Cardiology) UK, Interventional Fellow, New York, USA. Consultant Interventional Cardiologist.



Dr Rahoul D Sawant

MD (Medicine) CCT, Cardiology (UK), MRCP (UK), FACC (USA) Interventional Fellow, New York, USA Consultant Interventional Cardiologist

MMC Reg. No.: 85360 Asst. Professor, Cardiology, BVMCH, Pune

2998: Mrs.Bhutada Shobha (60y, Female)

Date: 03-Feb-2021

BP 159 / 119 mmHg Pulse 88 bpm Height 158 cm Weight 64.9 kg Temperature 97.2 F SPO2 98 % BMI 26.00 Kg/m²

Complaints: SOB ON EXERTION CLASS 3, PND

Diagnosis: DM, HTN

Sys.Exam: CVS: ECG: SR. T wave inversion in lateral leads.

2DECHO:

AO 20 LA 36 IVS 14 PW 14 LV D42 LV S 28

E/A 1.49

Normal size LV with moderate to severely impaired LV function, Anterior wall, lateral wall, apex and distal steptum are severely hypokinetic.

Rheumatic involvement of mitral valve. AML mildly thickened and PML movement is restricted. moderate to severe MR

Aortic valve is normal

Mild TR. Mild PH.; General: F 203 PP 336 Lipid Profile: TC-173, TRG-115, HDL-19, LDL-130, NonHDL-153, TC/HDL Ratio:9.0, BSL-183mg.;



@ www.emedica.in info.emedica@gmail.com

Heart - AFTER



Dr Rahoul D Sawant

MD (Medicine) CCT, Cardiology (UK), MRCP (UK), FACC (USA) Interventional Fellow, New York, USA Consultant Interventional Cardiologist MMC Reg. No.: 85360 Asst. Professor, Cardiology, BVMCH, Pune

2998 : Mrs.Bhutada Shobha (61y, Female)

Date: 01-May-2021

BP 160 / 84 mmHg Pulse 67 bpm Height 160 cm Weight 58.7 kg Temperature 97.2 F SPO2 98 % BMI 22.93 Kg/m²

Complaints: FLUCTUATING BSL, NO CHEST PAIN OR SOB

Diagnosis: DM, HTN, EF 30%, CAG 10.02.20201 TVD FOR SURGICAL OPINION, CREAT 1.28

Sys.Exam: CVS: ECG: SR. ST T changes in lateral leads.

2DECHO:

Ao 20 LA 37 IVS 12 PW 12 LV D 41 LV S 29

Normal LV size and mildly impaired LV systolic function.

EF 45%. EF has improved significantly.

Rheumatic involvement of mitral valve. AML thickened. PML movement restricted.

moderate MR

Normal aortic valve.

Mild TR. RVSP 15 + 10 mm of Hg.;



Heart - BEFORE







s*erlina





ATIENT'S NAME

Mr. Bobby Singh Poonia (38 Years/Male)

REPORT OF CORONARY ANGIOGRAPHY

ATE OF ANGIO

8th July, 2022

INGIO NO.

E-12448

P No

: 1217815

AO PRESSURE

: 120/80 mm of Hg

Procedure was done via Right Radial Artery.

LMCA

: Normal

LAD

: 99% long lesion in mid part

DIAGONAL

: 75% lesion

LCX

: 40% proximal lesion, OM1 - 80% lesion OM2 - 75% - 80% multiple lesions

OM3 - 100% lesion

RCA

: 85% mid lesion followed by 100% lesion. Distal part in filled by collaterals from LCA.

LV Angio

: Not done

FINAL DIAGNOSIS

> CORONARY ARTERY DISEASE - CRITICAL TRIPLE VESSEL DISEASE

> ACUTE CORONARY SYNDROME

RECOMMENDATION : EARLY REVASCULARLIZATION



Dr. Rasesh Pothiwala MD, DNB (Cardiologist) Dr. Akash Ramchandani MD, DM (Cardiologist)

Dr. Rahul Joshi (Cardiac Associate)


Head of the Department

Prepared by: Sagar Prajapati



Heart - AFTER

DD1 Job ID - 8723186 K953598/418 675 7658



This gentleman does have a surgical pattern of coronary artery disease. However, he is having minimal symptoms at present. Therefore, after discussion with him, we are opting to treat him medically in the first instance. We have started him on Bisoprolol 1.25 mg once daily which I would be grateful if you could continue. He will have a routine follow up with Dr Khan and if his symptoms become more intrusive then he will be referred for coronary artery bypass grafting.

Yours sincerely



Electronically approved

Dr Ashwin Radhakrishnan ST7 Cardiology

ICC: Dr Paul Jordan Consultant Cardiologist QEH



Eyes Refraction Error -BEFORE



Name: Mr. RAJKUMAR HAEMDEV

MR No: 00001010/PUNE

Ref by :

Date : 16/Apr/2015

Age/Sex : /Male

VISIT REPORT

Complaints: FOR RETINAL EVALUATION & MANAGEMENT USING SPECTS

SINCE THE AGE OF 20YRS -

: DM SINCE 45YRS ON RX, History **BORN WITH HERNIA-NOT OPERATED**

On Examination **BCVA**

Right Eye 6/9 N/6

Left Eye 6/6 N/6

K'metry (AutoK)

KH: 41.25/8.20X1 KV: 40.75/8.26X91

KH: 41.00/8.22X4 KV: 41.25/8.17X94

Autoref IOP

+0.25/-0.50X70 17mmHg

-0.25/-1.00X13 15mmHg

Slit Lamp Exam.

NS₁

NS₁

Fundus

MODERATE NPDR

MODERATE NPDR

Advice

TAB MACUGOLD ONCE A DAY FOR 3 MONTH EYE DROPS REFRESH TEARS 3 TIMES A DAY A MONTH

CHARLES OF	N. County	Right	Eye		Left Eye			
	Sph	Cyl	Axis	VA	Sph	Cyl	Axis	VA
Dist	-0.50	-0.75	70	6/9	0.00	-1.00	10	6/6
Near Add	+2.50			N/6	+2.50	W		N/6

Dr VARDHAMAN Kankariya



Eyes Refraction Error -AFTER

Or. Agarwals
Eye Hospital

For eyes like new

31/1, Kutika, Pune - Solapur Road, Hadapsar, Pune - 411013

Call: 7338987777 GST No.: 33AAACD2373G1Z2

Patient:

MR. RAJKUMAR HEMDEV

Age/Sex: Contact:

79 Years /Male 9823338118

MR No.:

HDP/4725/20

Doctor:

Facility:

Dr. MAHESH PATHAK

Dr Agarwal's Eye Hospital-HDP

Appt. Dt : Note Dt:

24 Nov'20 24 Nov'20

OPD SUMMARY

HISTORY

WANT FULL EYE CHECKUP

Known cases of OU CAT SX DONE 1 YEAR BACK, Diabetes since 40 years - on medication, Hypertension since 40 years - ON MEDICATION .

ASSES PRESCRIPTIONS

	R/OD	The second		
	Sph	Cyl	Axis	Vision
Distant	+0.00			6/6-2
Near	+1.75		/	N6

	L/OS			
	Sph	Cyl	Axis	Vision
Distant	0.00			6/6P
Near Near	+1.75	3Y		N6

EFRACTION



:31/01/2018

Leukaemia - BEFORE

Patient : Mrs. RUKMINI SHROFF :867007 IP No. Age/Sex : 62 Yrs/Female Visit No. : OP-2 Date Referred By : Dr. Shashikant Apte Sample Collected : 31/01/2018 11:00 Location : OPD Sample Revd. in Lab : 31/01/2018 11:53 Reported On Sponsor : 31/01/2018 15:56 Collected At : Sahyadri Speciality Hospitals Processed At : SSL Main Lab Lab No. :0031503118 : Verified Status

HAEMOGRAM Specimen : EDTA Whole Blood					
Test Name	Test Value	UOM	Biological Reference Interval		
Haemoglobin	10.9	g/dL	12.0 - 15.0		
R.B.C. Count	3.26	x 10^6/ul	3.80 - 4.80		
Haematocrit	31.6	76	36.0 - 46.0		
M.C,V.	96.9	n	76 - 96		
M.C.H	33.6	PE	27 - 32		
M.C.H.C.	34.7	gma/dl	31.5 - 34.5		
R.D.WCV	18.7	74	11.6-14		
Total W.B.C. Count	175600	/µL	4000 - 10000		
Differential Count					
Blast	5	76			
Myelocytes	8	26			
Meta+Band	9	96			
Neutrophils	63.0	26	40.0 - 80.0		
Lymphocytes	6.0	56	20.0 - 40.0		
Monocytes	2.0	96	2.0 - 10.0		
Eosinophils	4.0	%	1.0 - 6.0		
Basophils	3.0	96	<1.7		

Smear Study

Marked neutrophilic leucocytosis with shift to left upto blasts (5%).

Normocytic normochromic anemia, anisocytosis +.

Platelets adequate on smear.

10.3

Performed on Beckman Coulter Haematology Analyzer.

Platelet Count

MPV

End Of Report-

/µL

fl

150000 - 410000



Leukaemia - AFTER

Sahyadri Speciality Labs



Patient Mrs. RUKMINI SHROFF : 867007 IP No. Age/Sex : 63 Yrs/Female Visit No. : OP-8 Date Sample Collected Referred By : Dr. Shashikant Apte : 26/10/2018 11:59 : OPD Location Sample Revd. in Lab-: 26/10/2018 12:46 Sponsor Reported On : 26/10/2018 14:35 Collected At : Sahyadri Speciality Labs Processed At : SSL Main Lab Lab No. : 0057229918 Status : Verified HAEMOGRAM

Specimen : EDTA Whole Blood UOM Test Name Test Value Biological Reference Interval Haemoglobin g/dL 12.0 - 15.0 R.B.C. Count 3.24 x 10^6/ul 3.80 - 4.80 Haematocrit 31.3 96 36.0 - 46.0 MCV. 96.5 n 76-96 M.C.H 31 pg 27 - 32 M.C.H.C. 32.1 31.5 - 34.5 R.D.W.-CV 11.5 - 14 16.6 Total W.B.C. Count 7300 /µL 4000 - 10000 Differential Count Neutrophils 69.7 40.0 - 80.0 % Lymphocytes 21.5 20.0 - 40.0 Monocytes 3.9 96 2.0 - 10.0 Eosinophils 4.5 96 1.0 - 6.0 Basophils 0.4 <1-2 Platelet Count 323000 150000 - 410000 /µL

Smear Study Normocytic normochromic RBCs. Platelets adequate on smear.

Performed on Beckman Coulter Haematology Analyzer.

-End Of Report-

1 Real

Dr. Rajesh Phatale M.B.B.S., M.D. (Pathology) Reg. No: 67843

Entered By:70000971

MPV

Page 1 of 1 Sahyadri Hospitals Ltd. CIN: U85110PN1998PLC099499 Regd. Office: Plot No. 54, S.No. 89-90, Lokmanya Colony, Nr. R.K. Auto S. Jeet. Opp. Vanaz Company, Koftrud. Pune 411038.

